

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040523\
 Data File : VY013201.D
 Acq On : 05 Apr 2023 10:03
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 06 01:25:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 03 13:40:57 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.490	0.444	9.4	85	0.00
3 P	Chloromethane	0.511	0.422	17.4	81	0.00
4 C	Vinyl Chloride	0.512	0.496	3.1#	93	0.00
5 T	Bromomethane	0.362	0.333	8.0	91	0.00
6 T	Chloroethane	0.321	0.310	3.4	93	0.00
7 T	Trichlorofluoromethane	0.912	0.859	5.8	88	0.00
8 T	Diethyl Ether	0.286	0.313	-9.4	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.503	0.497	1.2	95	0.00
10 T	Methyl Iodide	0.655	0.700	-6.9	94	0.00
11 T	Tert butyl alcohol	0.048	0.050	-4.2	101	0.00
12 CM	1,1-Dichloroethene	0.486	0.484	0.4#	93	0.00
13 T	Acrolein	0.068	0.069	-1.5	95	0.00
14 T	Allyl chloride	0.828	0.734	11.4	84	0.00
15 T	Acrylonitrile	0.142	0.135	4.9	86	-0.01
16 T	Acetone	0.164	0.185	-12.8	90	0.00
17 T	Carbon Disulfide	1.619	1.425	12.0	83	0.00
18 T	Methyl Acetate	0.518	0.500	3.5	87	0.00
19 T	Methyl tert-butyl Ether	1.232	1.305	-5.9	95	0.00
20 T	Methylene Chloride	0.690	0.540	21.7	86	0.00
21 T	trans-1,2-Dichloroethene	0.548	0.509	7.1	87	-0.01
22 T	Diisopropyl ether	1.618	1.376	15.0	76	0.00
23 T	Vinyl Acetate	0.920	0.844	8.3	80	-0.01
24 P	1,1-Dichloroethane	0.958	0.829	13.5	82	0.00
25 T	2-Butanone	0.205	0.200	2.4	83	0.00
26 T	2,2-Dichloropropane	0.862	0.843	2.2	93	-0.01
27 T	cis-1,2-Dichloroethene	0.590	0.569	3.6	88	-0.01
28 T	Bromochloromethane	0.411	0.331	19.5	72	0.00
29 T	Tetrahydrofuran	0.122	0.115	5.7	81	0.00
30 C	Chloroform	0.992	0.971	2.1#	93	0.00
31 T	Cyclohexane	0.877	0.716	18.4	77	0.00
32 T	1,1,1-Trichloroethane	0.877	0.892	-1.7	94	0.00
33 S	1,2-Dichloroethane-d4	0.537	0.526	2.0	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.327	0.325	0.6	87	0.00
36 T	1,1-Dichloropropene	0.492	0.487	1.0	85	0.00
37 T	Ethyl Acetate	0.279	0.266	4.7	80	0.00
38 T	Carbon Tetrachloride	0.543	0.599	-10.3	97	0.00
39 T	Methylcyclohexane	0.577	0.631	-9.4	91	0.00
40 TM	Benzene	1.420	1.388	2.3	85	0.00
41 T	Methacrylonitrile	0.124	0.136	-9.7	94	0.00
42 TM	1,2-Dichloroethane	0.435	0.470	-8.0	96	0.00
43 T	Isopropyl Acetate	0.500	0.503	-0.6	83	0.00
44 TM	Trichloroethene	0.391	0.405	-3.6	91	0.00
45 C	1,2-Dichloropropane	0.356	0.385	-8.1#	95	0.00
46 T	Dibromomethane	0.204	0.219	-7.4	93	0.00
47 T	Bromodichloromethane	0.493	0.545	-10.5	97	0.00
48 T	Methyl methacrylate	0.233	0.298	-27.9#	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040523\
 Data File : VY013201.D
 Acq On : 05 Apr 2023 10:03
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 06 01:25:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 03 13:40:57 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	93	0.00
50 S	Toluene-d8	1.233	1.221	1.0	86	0.00
51 T	4-Methyl-2-Pentanone	0.270	0.276	-2.2	83	0.00
52 CM	Toluene	0.881	0.906	-2.8#	88	0.00
53 T	t-1,3-Dichloropropene	0.504	0.536	-6.3	91	0.00
54 T	cis-1,3-Dichloropropene	0.575	0.571	0.7	86	0.00
55 T	1,1,2-Trichloroethane	0.282	0.298	-5.7	92	0.00
56 T	Ethyl methacrylate	0.355	0.399	-12.4	92	0.00
57 T	1,3-Dichloropropane	0.477	0.484	-1.5	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.166	0.189	-13.9	98	0.00
59 T	2-Hexanone	0.195	0.213	-9.2	85	0.00
60 T	Dibromochloromethane	0.355	0.400	-12.7	98	0.00
61 T	1,2-Dibromoethane	0.274	0.292	-6.6	94	0.00
62 S	4-Bromofluorobenzene	0.401	0.423	-5.5	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.384	0.415	-8.1	95	0.00
65 PM	Chlorobenzene	1.061	1.101	-3.8	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.433	-8.8	97	0.00
67 C	Ethyl Benzene	1.833	1.918	-4.6#	90	0.00
68 T	m/p-Xylenes	0.709	0.765	-7.9	91	0.00
69 T	o-Xylene	0.658	0.713	-8.4	92	0.00
70 T	Styrene	1.100	1.235	-12.3	95	0.00
71 P	Bromoform	0.247	0.301	-21.9	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.451	3.626	-5.1	94	0.00
74 T	N-amyl acetate	0.948	0.942	0.6	86	0.00
75 P	1,1,2,2-Tetrachloroethane	0.723	0.707	2.2	88	0.00
76 T	1,2,3-Trichloropropane	0.562	0.549	2.3	87	0.00
77 T	Bromobenzene	0.858	0.925	-7.8	98	0.00
78 T	n-propylbenzene	4.244	4.296	-1.2	90	0.00
79 T	2-Chlorotoluene	2.408	2.406	0.1	92	0.00
80 T	1,3,5-Trimethylbenzene	2.886	3.071	-6.4	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.252	0.266	-5.6	92	0.00
82 T	4-Chlorotoluene	2.506	2.533	-1.1	92	0.00
83 T	tert-Butylbenzene	2.478	2.733	-10.3	97	0.00
84 T	1,2,4-Trimethylbenzene	2.826	3.031	-7.3	95	0.00
85 T	sec-Butylbenzene	3.742	3.927	-4.9	93	0.00
86 T	p-Isopropyltoluene	3.081	3.372	-9.4	97	0.00
87 T	1,3-Dichlorobenzene	1.701	1.780	-4.6	98	0.00
88 T	1,4-Dichlorobenzene	1.705	1.790	-5.0	99	0.00
89 T	n-Butylbenzene	2.880	2.997	-4.1	93	0.00
90 T	Hexachloroethane	0.642	0.641	0.2	94	0.00
91 T	1,2-Dichlorobenzene	1.519	1.639	-7.9	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.130	-8.3	99	0.00
93 T	1,2,4-Trichlorobenzene	0.886	0.971	-9.6	97	0.00
94 T	Hexachlorobutadiene	0.545	0.624	-14.5	103	0.00
95 T	Naphthalene	1.657	2.095	-26.4#	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040523\
Data File : VY013201.D
Acq On : 05 Apr 2023 10:03
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 06 01:25:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 03 13:40:57 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.781	0.923	-18.2	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6