



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040523\
Data File : VY013200.D
Acq On : 05 Apr 2023 09:30
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Title : SW846 8260
Last Update : Mon Apr 03 13:40:57 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	11586	PASS
75	95	30	60	52.3	32085	PASS
95	95	100	100	100.0	61384	PASS
96	95	5	9	6.8	4171	PASS
173	174	0.00	2	1.5	883	PASS
174	95	50	100	95.9	58877	PASS
175	174	5	9	7.6	4469	PASS
176	174	95	101	97.0	57123	PASS
177	176	5	9	6.8	3901	PASS