

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
 Data File : VY013439.D
 Acq On : 25 Apr 2023 12:23
 Operator : KP/MD
 Sample : 02406-05RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-14-8.0-8.5

Quant Time: Apr 25 15:12:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	18184	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	27623	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	23938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	8679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	11353	56.954	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.900%
35) Dibromofluoromethane	7.716	113	9793	56.367	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.740%
50) Toluene-d8	10.179	98	26259	42.487	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.980%
62) 4-Bromofluorobenzene	12.477	95	7517	36.077	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	72.160%
Target Compounds						
16) Acetone	3.954	43	1437	18.306	ug/l	# 65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
Data File : VY013439.D
Acq On : 25 Apr 2023 12:23
Operator : KP/MD
Sample : 02406-05RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-14-8.0-8.5

Quant Time: Apr 25 15:12:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

