

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
 Data File : VY013447.D
 Acq On : 25 Apr 2023 15:29
 Operator : KP/MD
 Sample : 02477-03
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 W-H4(ELEV-22)

Quant Time: Apr 26 02:50:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.789	168	160598	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	272830	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	261653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	122441	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	109918	62.435	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.880%
35) Dibromofluoromethane	7.716	113	96776	56.397	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.800%
50) Toluene-d8	10.179	98	337814	55.339	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.680%
62) 4-Bromofluorobenzene	12.477	95	107990	52.474	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.940%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
 Data File : VY013447.D
 Acq On : 25 Apr 2023 15:29
 Operator : KP/MD
 Sample : 02477-03
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 W-H4(ELEV-22)

Quant Time: Apr 26 02:50:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

