

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
 Data File : VY013454.D
 Acq On : 25 Apr 2023 18:13
 Operator : KP/MD
 Sample : 02389-30RE
 Misc : 3.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SPB-8WC-23-04-17

Quant Time: Apr 26 02:53:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	163062	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	271813	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	249314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	100839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	108314	60.595	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.180%
35) Dibromofluoromethane	7.716	113	96034	56.174	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.340%
50) Toluene-d8	10.179	98	327149	53.793	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.580%
62) 4-Bromofluorobenzene	12.477	95	96245	46.942	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.880%
Target Compounds						Qvalue
20) Methylene Chloride	4.716	84	34597	9.065	ug/l	97
25) 2-Butanone	6.972	43	4794	5.370	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
Data File : VY013454.D
Acq On : 25 Apr 2023 18:13
Operator : KP/MD
Sample : 02389-30RE
Misc : 3.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SPB-8WC-23-04-17

Quant Time: Apr 26 02:53:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

