

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
 Data File : VY013456.D
 Acq On : 25 Apr 2023 19:00
 Operator : KP/MD
 Sample : 02470-03
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-04-9.5-10.0

Quant Time: Apr 26 02:53:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	4828	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.697	114	8229	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	9281	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	3550	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	5635	106.470	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	212.940%#
35) Dibromofluoromethane	7.722	113	4084	78.908	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	157.820%#
50) Toluene-d8	10.179	98	10396	56.463	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.920%
62) 4-Bromofluorobenzene	12.483	95	3385	54.534	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
Data File : VY013456.D
Acq On : 25 Apr 2023 19:00
Operator : KP/MD
Sample : 02470-03
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-04-9.5-10.0

Quant Time: Apr 26 02:53:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

