

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
 Data File : VY013457.D
 Acq On : 25 Apr 2023 19:23
 Operator : KP/MD
 Sample : 02470-05
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-06-9.5-10.0

Quant Time: Apr 26 02:54:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	25146	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	37412	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	33906	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	11753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	16789	60.906	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 121.820%		
35) Dibromofluoromethane	7.716	113	14385	61.134	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 122.260%		
50) Toluene-d8	10.179	98	43703	52.209	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 104.420%		
62) 4-Bromofluorobenzene	12.477	95	11192	39.660	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 79.320%		
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	4678	6.760	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
Data File : VY013457.D
Acq On : 25 Apr 2023 19:23
Operator : KP/MD
Sample : 02470-05
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-06-9.5-10.0

Quant Time: Apr 26 02:54:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

