

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
 Data File : VY013451.D
 Acq On : 25 Apr 2023 17:03
 Operator : KP/MD
 Sample : 02390-06
 Misc : 3.13g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SPC-2WC-23-04-17

Quant Time: Apr 26 02:52:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	160363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	273019	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	246387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	94893	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	117139	66.635	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.260%
35) Dibromofluoromethane	7.716	113	96041	55.930	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.860%
50) Toluene-d8	10.179	98	326388	53.430	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.860%
62) 4-Bromofluorobenzene	12.477	95	91508	44.435	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.860%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042523\
Data File : VY013451.D
Acq On : 25 Apr 2023 17:03
Operator : KP/MD
Sample : 02390-06
Misc : 3.13g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SPC-2WC-23-04-17

Quant Time: Apr 26 02:52:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY013451.D
Acq: 25 Apr 2023 17:03




Instrument :
MSVOA_Y
ClientSampleId :
SPC-2WC-23-04-17


Tgt Ion:168 Resp: 160363
Ion Ratio Lower Upper
168 100
99 53.3 41.8 62.6



#33
1,2-Dichloroethane-d4
Concen: 66.635 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY013451.D
Acq: 25 Apr 2023 17:03

Tgt Ion: 65 Resp: 117139
Ion Ratio Lower Upper
65 100
67 50.4 0.0 103.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. -0.006 min

Lab File: VY013451.D

Acq: 25 Apr 2023 17:03

Instrument :

MSVOA_Y

ClientSampleId :

SPC-2WC-23-04-17





Tgt Ion:114 Resp: 273019

Ion Ratio Lower Upper

114 100

63 19.8 0.0 36.6

88 15.1 0.0 28.8



#35

Dibromofluoromethane

Concen: 55.930 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY013451.D

Acq: 25 Apr 2023 17:03




Tgt Ion:113 Resp: 96041


Ion Ratio Lower Upper

113 100

111 101.2 82.2 123.2

192 20.2 16.7 25.1









#50
Toluene-d8
Concen: 53.430 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY013451.D
Acq: 25 Apr 2023 17:03




Instrument :
MSVOA_Y
ClientSampleId :
SPC-2WC-23-04-17


Tgt Ion: 98 Resp: 326388
Ion Ratio Lower Upper
98 100
100 63.5 50.5 75.7



#62
4-Bromofluorobenzene
Concen: 44.435 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.000 min
Lab File: VY013451.D
Acq: 25 Apr 2023 17:03

Tgt Ion: 95 Resp: 91508
Ion Ratio Lower Upper
95 100
174 91.4 0.0 182.0
176 88.7 0.0 175.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY013451.D

Acq: 25 Apr 2023 17:03

Instrument :

MSVOA_Y

ClientSampleId :

SPC-2WC-23-04-17





Tgt Ion:117 Resp: 246387

Ion Ratio Lower Upper

117 100

82 53.4 43.0 64.6

119 30.5 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. -0.000 min

Lab File: VY013451.D

Acq: 25 Apr 2023 17:03

Tgt Ion:152 Resp: 94893

Ion Ratio Lower Upper

152 100

115 56.3 28.1 84.2

150 159.9 0.0 352.0

