

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042823\
 Data File : VY013525.D
 Acq On : 28 Apr 2023 14:54
 Operator : KP/MD
 Sample : 02502-10RE
 Misc : 5.92g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-511-VOC-02

Quant Time: Apr 28 23:13:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	209834	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	350612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	346561	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	154836	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	145414	63.217	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.440%
35) Dibromofluoromethane	7.716	113	122972	55.765	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.520%
50) Toluene-d8	10.179	98	435789	55.552	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	111.100%
62) 4-Bromofluorobenzene	12.483	95	137772	52.094	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042823\
Data File : VY013525.D
Acq On : 28 Apr 2023 14:54
Operator : KP/MD
Sample : 02502-10RE
Misc : 5.92g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-511-VOC-02

Quant Time: Apr 28 23:13:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

