

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050223\
 Data File : VY013560.D
 Acq On : 02 May 2023 16:34
 Operator : KP/MD
 Sample : 02568-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11938

Quant Time: May 03 01:34:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	4790	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	8841	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	9575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	6149	117.104	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 234.200%#	
35) Dibromofluoromethane	7.722	113	3979	71.557	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 143.120%	
50) Toluene-d8	10.179	98	10663	53.904	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 107.800%	
62) 4-Bromofluorobenzene	12.483	95	3423	51.329	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 102.660%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050223\
Data File : VY013560.D
Acq On : 02 May 2023 16:34
Operator : KP/MD
Sample : 02568-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11938

Quant Time: May 03 01:34:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

