

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050223\
 Data File : VY013565.D
 Acq On : 02 May 2023 18:31
 Operator : KP/MD
 Sample : 02583-07
 Misc : 12.32g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CF-27

Quant Time: May 03 01:35:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	197160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	339655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	330154	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	148958	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	150899	69.818	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.640%
35) Dibromofluoromethane	7.710	113	118922	55.668	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.340%
50) Toluene-d8	10.179	98	410218	53.979	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.960%
62) 4-Bromofluorobenzene	12.477	95	130749	51.034	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050223\
Data File : VY013565.D
Acq On : 02 May 2023 18:31
Operator : KP/MD
Sample : 02583-07
Misc : 12.32g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CF-27

Quant Time: May 03 01:35:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

