

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050823\
 Data File : VY013646.D
 Acq On : 08 May 2023 14:02
 Operator : KP/MD
 Sample : 02622-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 2264

Quant Time: May 09 04:56:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	173084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	280134	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	280907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	133168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	105530	55.619	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.240%
35) Dibromofluoromethane	7.716	113	93403	53.012	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.020%
50) Toluene-d8	10.179	98	352490	56.238	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.480%
62) 4-Bromofluorobenzene	12.477	95	123943	58.656	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	117.320%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.716	84	21561	1.338	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050823\
Data File : VY013646.D
Acq On : 08 May 2023 14:02
Operator : KP/MD
Sample : 02622-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
2264

Quant Time: May 09 04:56:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

