

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050823\
 Data File : VY013648.D
 Acq On : 08 May 2023 14:49
 Operator : KP/MD
 Sample : 02616-01RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-1RE

Quant Time: May 09 04:57:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	41206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	64927	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	68163	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	30501	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	27176	60.163	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.320%
35) Dibromofluoromethane	7.716	113	22224	54.422	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.840%
50) Toluene-d8	10.179	98	84003	57.825	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	115.660%
62) 4-Bromofluorobenzene	12.477	95	33389	68.176	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	136.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050823\
Data File : VY013648.D
Acq On : 08 May 2023 14:49
Operator : KP/MD
Sample : 02616-01RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-1RE

Quant Time: May 09 04:57:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

