

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050922\
 Data File : VY008667.D
 Acq On : 09 May 2022 17:50
 Operator : KP/MD
 Sample : N2619-17RE
 Misc : 5.72g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SS08-1-2-FTRE

Quant Time: May 10 03:24:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	158572	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	242082	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	208379	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	92495	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	69371	51.056	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.120%
35) Dibromofluoromethane	7.716	113	72678	48.718	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.440%
50) Toluene-d8	10.179	98	275409	52.388	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.780%
62) 4-Bromofluorobenzene	12.483	95	82428	43.582	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050922\
Data File : VY008667.D
Acq On : 09 May 2022 17:50
Operator : KP/MD
Sample : N2619-17RE
Misc : 5.72g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SS08-1-2-FTRE

Quant Time: May 10 03:24:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration

