



Tert butyl alcohol



Response = $3.530 \times 10^{-2} \times \text{Amt} + 8.600 \times 10^{-3}$
 Coef of Det (r^2) = 0.997583 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y050922S.M
 Calibration Table Last Updated: Mon May 09 19:21:20 2022

Methyl Acetate

Response Ratio



$$\text{Response} = 2.909\text{e-}001 * \text{Amt} + 9.144\text{e-}003$$


Coef of Det (r^2) = 0.994417 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M

Calibration Table Last Updated: Mon May 09 19:21:20 2022

Methacrylonitrile

Response Ratio



$$\text{Response} = 1.379\text{e-}001 * \text{Amt} - 2.116\text{e-}003$$

Coef of Det (r^2) = 0.994688 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M

Calibration Table Last Updated: Mon May 09 19:21:20 2022