

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050923\
 Data File : VY013667.D
 Acq On : 09 May 2023 14:48
 Operator : KP/MD
 Sample : 02665-10
 Misc : 5.80g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-612-VOC-02

Quant Time: May 10 05:37:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	200938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	338000	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	326386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	152657	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	145881	66.228	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.460%
35) Dibromofluoromethane	7.716	113	118705	55.838	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.680%
50) Toluene-d8	10.179	98	415638	54.960	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.920%
62) 4-Bromofluorobenzene	12.483	95	136703	53.619	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	107.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050923\
Data File : VY013667.D
Acq On : 09 May 2023 14:48
Operator : KP/MD
Sample : 02665-10
Misc : 5.80g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-612-VOC-02

Quant Time: May 10 05:37:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

