

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050923\
 Data File : VY013675.D
 Acq On : 09 May 2023 17:55
 Operator : KP/MD
 Sample : 02624-03
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WZ-2B

Quant Time: May 10 05:40:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	73225	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	106935	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	97261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	36899	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	41687	51.933	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.860%
35) Dibromofluoromethane	7.716	113	37738	56.110	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.220%
50) Toluene-d8	10.179	98	129495	54.123	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.240%
62) 4-Bromofluorobenzene	12.477	95	33920	42.053	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050923\
Data File : VY013675.D
Acq On : 09 May 2023 17:55
Operator : KP/MD
Sample : 02624-03
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WZ-2B

Quant Time: May 10 05:40:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

