

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051023\
 Data File : VY013684.D
 Acq On : 10 May 2023 13:46
 Operator : KP/MD
 Sample : 02723-01
 Misc : 5.36g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-14-G1-SO-7.5-050923

Quant Time: May 11 03:26:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	196861	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	335058	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	326671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	149384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	140779	65.235	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.460%
35) Dibromofluoromethane	7.716	113	117117	55.575	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.160%
50) Toluene-d8	10.179	98	412478	55.021	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.040%
62) 4-Bromofluorobenzene	12.483	95	130643	51.692	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.380%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	3.893	101	20416	9.619	ug/l	97
16) Acetone	3.954	43	9745	11.467	ug/l #	87
44) Trichloroethene	8.941	130	27829	10.619	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051023\
Data File : VY013684.D
Acq On : 10 May 2023 13:46
Operator : KP/MD
Sample : 02723-01
Misc : 5.36g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-7.5-050923

Quant Time: May 11 03:26:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

