

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051023\
 Data File : VY013692.D
 Acq On : 10 May 2023 16:53
 Operator : KP/MD
 Sample : 02721-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: May 11 03:30:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	17963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	27387	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	24566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	8971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	12233	62.124	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.240%
35) Dibromofluoromethane	7.722	113	10525	61.102	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.200%
50) Toluene-d8	10.179	98	32842	53.596	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.200%
62) 4-Bromofluorobenzene	12.483	95	9532	46.142	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.280%
Target Compounds						
20) Methylene Chloride	4.716	84	3780	8.912	ug/l	97
44) Trichloroethene	8.941	130	3681	17.185	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051023\
Data File : VY013692.D
Acq On : 10 May 2023 16:53
Operator : KP/MD
Sample : 02721-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: May 11 03:30:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

