

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051322\
 Data File : VY008743.D
 Acq On : 13 May 2022 12:48
 Operator : KP/MD
 Sample : VY0513SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0513SBL01

Quant Time: May 16 05:50:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051222S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 13 02:08:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	167154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	293628	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	261298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	103357	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	98614	59.239	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.480%
35) Dibromofluoromethane	7.716	113	94521	53.634	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.260%
50) Toluene-d8	10.179	98	352581	53.050	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.100%
62) 4-Bromofluorobenzene	12.483	95	98416	42.740	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051322\
Data File : VY008743.D
Acq On : 13 May 2022 12:48
Operator : KP/MD
Sample : VY0513SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBL01

Quant Time: May 16 05:50:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051222S.M
Quant Title : SW846 8260
QLast Update : Fri May 13 02:08:33 2022
Response via : Initial Calibration

