

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051322\
 Data File : VY008754.D
 Acq On : 13 May 2022 17:11
 Operator : KP/MD
 Sample : N2781-19
 Misc : 5.20g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P018-SS001-0006-01MSD

Quant Time: May 16 05:53:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051222S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 13 02:08:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	143350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	258220	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	230211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	94641	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	91400	64.023	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.040%
35) Dibromofluoromethane	7.716	113	81818	52.792	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.580%
50) Toluene-d8	10.179	98	306849	52.500	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.000%
62) 4-Bromofluorobenzene	12.483	95	85738	42.340	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051322\
Data File : VY008754.D
Acq On : 13 May 2022 17:11
Operator : KP/MD
Sample : N2781-19
Misc : 5.20g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P018-SS001-0006-01MSD

Quant Time: May 16 05:53:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051222S.M
Quant Title : SW846 8260
QLast Update : Fri May 13 02:08:33 2022
Response via : Initial Calibration

