

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051322\
 Data File : VY008761.D
 Acq On : 13 May 2022 19:55
 Operator : KP/MD
 Sample : N2775-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-35S

Quant Time: May 16 05:55:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051222S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 13 02:08:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	49446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	86039	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	75595	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	28603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	31033	63.020	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.040%	
35) Dibromofluoromethane	7.716	113	29881	57.864	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.720%	
50) Toluene-d8	10.179	98	101443	52.090	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.180%	
62) 4-Bromofluorobenzene	12.483	95	26949	39.940	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 79.880%	
Target Compounds						
16) Acetone	3.942	43	2403	18.363	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051322\
Data File : VY008761.D
Acq On : 13 May 2022 19:55
Operator : KP/MD
Sample : N2775-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-35S

Quant Time: May 16 05:55:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051222S.M
Quant Title : SW846 8260
QLast Update : Fri May 13 02:08:33 2022
Response via : Initial Calibration

