

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051322\
 Data File : VY008744.D
 Acq On : 13 May 2022 13:17
 Operator : KP/MD
 Sample : VY0513SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0513SBS01

Quant Time: May 16 05:50:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051222S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 13 02:08:33 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022
 Supervised By :Semsettin Yesilyurt 05/16/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	170393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	287281	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	271526	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	129997	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	82613	48.684	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	97.360%		
35) Dibromofluoromethane	7.716	113	81721	47.396	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	94.800%		
50) Toluene-d8	10.179	98	284305	43.723	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	87.440%		
62) 4-Bromofluorobenzene	12.483	95	99730	44.267	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	88.540%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	32616	23.885	ug/l	99
3) Chloromethane	2.113	50	61930	21.943	ug/l	96
4) Vinyl Chloride	2.247	62	83089	21.748	ug/l	98
5) Bromomethane	2.637	94	80759	22.565	ug/l	99
6) Chloroethane	2.790	64	55926	21.234	ug/l	97
7) Trichlorofluoromethane	3.119	101	66252	22.452	ug/l	98
8) Diethyl Ether	3.527	74	20609	22.658	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.899	101	39581	22.658	ug/l	99
10) Methyl Iodide	4.088	142	35745	21.251	ug/l	98
11) Tert butyl alcohol	4.948	59	22776	57.625	ug/l	97
12) 1,1-Dichloroethene	3.869	96	35197	22.001	ug/l	92
13) Acrolein	3.723	56	14595	133.877	ug/l	94
14) Allyl chloride	4.479	41	47885	21.186	ug/l	99
15) Acrylonitrile	5.155	53	50687	108.993	ug/l	100
16) Acetone	3.948	43	38833	99.659	ug/l	97
17) Carbon Disulfide	4.186	76	104014	22.854	ug/l	99
18) Methyl Acetate	4.479	43	25728	19.742	ug/l	96
19) Methyl tert-butyl Ether	5.216	73	101899	21.487	ug/l	96
20) Methylene Chloride	4.710	84	56677	23.859	ug/l	98
21) trans-1,2-Dichloroethene	5.216	96	41630	21.366	ug/l	96
22) Diisopropyl ether	6.112	45	114009	21.636	ug/l	95
23) Vinyl Acetate	6.057	43	368224	117.223	ug/l	100
24) 1,1-Dichloroethane	6.015	63	71720	21.527	ug/l	97
25) 2-Butanone	6.978	43	64736	107.550	ug/l	94
26) 2,2-Dichloropropane	6.978	77	68044	21.488	ug/l	100
27) cis-1,2-Dichloroethene	6.984	96	47573	21.426	ug/l	100
28) Bromochloromethane	7.332	49	26650	20.045	ug/l	94
29) Tetrahydrofuran	7.350	42	41555	107.056	ug/l	98
30) Chloroform	7.508	83	78446	22.090	ug/l	90
31) Cyclohexane	7.777	56	64765	21.128	ug/l	100
32) 1,1,1-Trichloroethane	7.697	97	68164	21.214	ug/l	99
36) 1,1-Dichloropropene	7.911	75	57492	20.854	ug/l	98
37) Ethyl Acetate	7.069	43	30511	21.870	ug/l	97
38) Carbon Tetrachloride	7.899	117	60831	20.910	ug/l	94
39) Methylcyclohexane	9.179	83	73617	21.003	ug/l	98
40) Benzene	8.155	78	165763	20.689	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051322\
 Data File : VY008744.D
 Acq On : 13 May 2022 13:17
 Operator : KP/MD
 Sample : VY0513SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0513SBS01

Quant Time: May 16 05:50:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051222S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 13 02:08:33 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022
 Supervised By :Semsettin Yesilyurt 05/16/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.319	41	15748m	23.268	ug/l	
42) 1,2-Dichloroethane	8.234	62	50295	22.183	ug/l	100
43) Isopropyl Acetate	8.271	43	56482	21.482	ug/l #	88
44) Trichloroethene	8.941	130	45712	21.365	ug/l	91
45) 1,2-Dichloropropane	9.215	63	41995	21.709	ug/l	98
46) Dibromomethane	9.301	93	26978	21.978	ug/l	98
47) Bromodichloromethane	9.496	83	61713	21.695	ug/l	99
48) Methyl methacrylate	9.295	41	24722	21.950	ug/l	97
49) 1,4-Dioxane	9.301	88	6901	448.798	ug/l	94
51) 4-Methyl-2-Pentanone	10.069	43	150338	106.635	ug/l	97
52) Toluene	10.240	92	109936	20.624	ug/l	98
53) t-1,3-Dichloropropene	10.465	75	61340	20.917	ug/l	96
54) cis-1,3-Dichloropropene	9.929	75	70590	21.544	ug/l	97
55) 1,1,2-Trichloroethane	10.642	97	34904	20.739	ug/l	98
56) Ethyl methacrylate	10.508	69	43906	20.269	ug/l	93
57) 1,3-Dichloropropane	10.788	76	59573	21.577	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.782	63	111961	98.694	ug/l	97
59) 2-Hexanone	10.831	43	101132	106.445	ug/l	96
60) Dibromochloromethane	10.983	129	41706	20.712	ug/l	98
61) 1,2-Dibromoethane	11.087	107	34907	21.563	ug/l	95
64) Tetrachloroethene	10.721	164	41267	20.239	ug/l	92
65) Chlorobenzene	11.514	112	117145	20.285	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.587	131	42930	20.020	ug/l	97
67) Ethyl Benzene	11.593	91	207661	20.028	ug/l	99
68) m/p-Xylenes	11.703	106	162974	39.419	ug/l	100
69) o-Xylene	12.032	106	77260	19.873	ug/l	99
70) Styrene	12.044	104	129508	20.010	ug/l	99
71) Bromoform	12.209	173	26125	20.701	ug/l #	97
73) Isopropylbenzene	12.331	105	199279	20.264	ug/l	100
74) N-amyl acetate	12.142	43	48751	21.444	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.581	83	42803	21.189	ug/l	100
76) 1,2,3-Trichloropropane	12.636	75	25153m	19.228	ug/l	
77) Bromobenzene	12.611	156	45758	20.550	ug/l	98
78) n-propylbenzene	12.672	91	247981	20.820	ug/l	98
79) 2-Chlorotoluene	12.758	91	133335	20.791	ug/l	100
80) 1,3,5-Trimethylbenzene	12.812	105	165374	20.623	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.380	75	14778	21.825	ug/l	97
82) 4-Chlorotoluene	12.855	91	140038	20.741	ug/l	100
83) tert-Butylbenzene	13.075	119	143964	20.207	ug/l	100
84) 1,2,4-Trimethylbenzene	13.123	105	165239	20.733	ug/l	100
85) sec-Butylbenzene	13.251	105	221586	20.498	ug/l	99
86) p-Isopropyltoluene	13.367	119	183582	20.347	ug/l	99
87) 1,3-Dichlorobenzene	13.367	146	94075	20.337	ug/l	98
88) 1,4-Dichlorobenzene	13.446	146	95282	20.987	ug/l	99
89) n-Butylbenzene	13.696	91	175511	20.798	ug/l	98
90) Hexachloroethane	13.959	117	34825	20.922	ug/l	98
91) 1,2-Dichlorobenzene	13.739	146	82296	20.540	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	6291	20.977	ug/l	90
93) 1,2,4-Trichlorobenzene	15.007	180	46703	20.991	ug/l	98
94) Hexachlorobutadiene	15.111	225	24417	20.845	ug/l	98
95) Naphthalene	15.239	128	97931	20.594	ug/l	99
96) 1,2,3-Trichlorobenzene	15.428	180	40640	20.630	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051322\
Data File : VY008744.D
Acq On : 13 May 2022 13:17
Operator : KP/MD
Sample : VY0513SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022

Quant Time: May 16 05:50:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051222S.M
Quant Title : SW846 8260
QLast Update : Fri May 13 02:08:33 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01

Tgt Ion:168 Resp: 17039
Ion Ratio Lower Upper
168 100
99 52.8 42.3 63.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#2
Dichlorodifluoromethane
Concen: 23.885 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion: 85 Resp: 32616
Ion Ratio Lower Upper
85 100
87 31.9 16.2 48.6





#3
Chloromethane
Concen: 21.943 ug/l
RT: 2.113 min Scan# 41
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion: 50 Resp: 61930
Ion Ratio Lower Upper
50 100
52 31.1 26.7 40.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#4
Vinyl Chloride
Concen: 21.748 ug/l
RT: 2.247 min Scan# 70
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion: 62 Resp: 83089
Ion Ratio Lower Upper
62 100
64 32.0 26.7 40.1





#5
Bromomethane
Concen: 22.565 ug/l
RT: 2.637 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion: 94 Resp: 80759
Ion Ratio Lower Upper
94 100
96 91.7 74.2 111.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#6
Chloroethane
Concen: 21.234 ug/l
RT: 2.790 min Scan# 159
Delta R.T. -0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion: 64 Resp: 55926
Ion Ratio Lower Upper
64 100
66 30.8 25.8 38.6





#7
Trichlorofluoromethane
Concen: 22.452 ug/l
RT: 3.119 min Scan# 214
Delta R.T. -0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:101 Resp: 6625
Ion Ratio Lower Upper
101 100
103 60.6 49.5 74.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#8
Diethyl Ether
Concen: 22.658 ug/l
RT: 3.527 min Scan# 280
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17



Tgt Ion: 74 Resp: 20609
Ion Ratio Lower Upper
74 100
45 83.0 42.6 127.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 22.658 ug/l
RT: 3.899 min Scan# 340
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01

Tgt Ion:101 Resp: 3958
Ion Ratio Lower Upper
101 100
85 45.7 37.8 56.6
151 74.6 59.4 89.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#10
Methyl Iodide
Concen: 21.251 ug/l
RT: 4.088 min Scan# 372
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion:142 Resp: 35745
Ion Ratio Lower Upper
142 100
127 46.0 35.4 53.0
141 14.5 11.7 17.5





#11
Tert butyl alcohol
Concen: 57.625 ug/l
RT: 4.948 min Scan# 513
Delta R.T. -0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion: 59 Resp: 22770
Ion Ratio Lower Upper
59 100
57 7.3 6.7 10.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#12
1,1-Dichloroethene
Concen: 22.001 ug/l
RT: 3.869 min Scan# 336
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion: 96 Resp: 35197
Ion Ratio Lower Upper
96 100
61 146.3 123.9 185.9
98 72.9 51.2 76.8





#13
Acrolein
Concen: 133.877 ug/l
RT: 3.723 min Scan# 313
Delta R.T. -0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 56 Resp: 14598
Ion Ratio Lower Upper
56 100
55 69.3 59.8 89.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#14
Allyl chloride
Concen: 21.186 ug/l
RT: 4.479 min Scan# 436
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion: 41 Resp: 47885
Ion Ratio Lower Upper
41 100
39 80.2 64.0 96.0
76 45.1 37.5 56.3





#15
Acrylonitrile
Concen: 108.993 ug/l
RT: 5.155 min Scan# 548
Delta R.T. -0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 53 Resp: 5068
Ion Ratio Lower Upper
53 100
52 82.7 65.8 98.6
51 36.8 29.6 44.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#16
Acetone
Concen: 99.659 ug/l
RT: 3.948 min Scan# 349
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion: 43 Resp: 38833
Ion Ratio Lower Upper
43 100
58 29.9 25.1 37.7





#17
Carbon Disulfide
Concen: 22.854 ug/l
RT: 4.186 min Scan# 388
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 76 Resp: 104014
Ion Ratio Lower Upper
76 100
78 9.0 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#18
Methyl Acetate
Concen: 19.742 ug/l
RT: 4.479 min Scan# 436
Delta R.T. 0.001 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Tgt Ion: 43 Resp: 25728
Ion Ratio Lower Upper
43 100
74 27.9 20.6 31.0





#19

Methyl tert-butyl Ether

Concen: 21.487 ug/l

RT: 5.216 min Scan# 51

Delta R.T. 0.000 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17

Instrument :

MSVOA_Y

ClientSampleId :

VY0513SBS01

Tgt Ion: 73 Resp: 10189

Ion Ratio Lower Upper

73 100





57 21.6 15.8 23.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022

Supervised By :Semsettin Yesilyurt 05/16/2022



#20

Methylene Chloride

Concen: 23.859 ug/l

RT: 4.710 min Scan# 474

Delta R.T. 0.006 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17

Tgt Ion: 84 Resp: 56677




Ion Ratio Lower Upper


84 100

49 111.3 88.6 133.0

51 31.2 27.4 41.0


86 64.5 50.2 75.2





#21
trans-1,2-Dichloroethene
Concen: 21.366 ug/l
RT: 5.216 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion: 96 Resp: 41630
Ion Ratio Lower Upper
96 100
61 131.1 101.8 152.6
98 58.7 50.2 75.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#22
Diisopropyl ether
Concen: 21.636 ug/l
RT: 6.112 min Scan# 704
Delta R.T. -0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion: 45 Resp: 114009
Ion Ratio Lower Upper
45 100
43 62.2 46.5 69.7
87 29.2 26.3 39.5
59 13.0 9.8 14.6





#23
 Vinyl Acetate
 Concen: 117.223 ug/l
 RT: 6.057 min Scan# 694
 Delta R.T. 0.006 min
 Lab File: VY008744.D
 Acq: 13 May 2022 13:17




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0513SBS01




Tgt Ion: 43 Resp: 368224
 Ion Ratio Lower Upper
 43 100
 86 12.9 10.2 15.2

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
 Supervised By :Semsettin Yesilyurt 05/16/2022




#24
 1,1-Dichloroethane
 Concen: 21.527 ug/l
 RT: 6.015 min Scan# 688
 Delta R.T. 0.006 min
 Lab File: VY008744.D
 Acq: 13 May 2022 13:17




Tgt Ion: 63 Resp: 71720
 Ion Ratio Lower Upper
 63 100
 98 7.9 3.3 9.9
 100 5.9 2.6 7.8





#25
2-Butanone
Concen: 107.550 ug/l
RT: 6.978 min Scan# 846
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 43 Resp: 64730
Ion Ratio Lower Upper
43 100
72 27.9 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#26
2,2-Dichloropropane
Concen: 21.488 ug/l
RT: 6.978 min Scan# 846
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion: 77 Resp: 68044
Ion Ratio Lower Upper
77 100
97 22.9 11.5 34.5





#27
cis-1,2-Dichloroethene
Concen: 21.426 ug/l
RT: 6.984 min Scan# 846
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 96 Resp: 4757
Ion Ratio Lower Upper
96 100
61 135.6 0.0 270.4
98 63.3 0.0 126.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#28
Bromochloromethane
Concen: 20.045 ug/l
RT: 7.332 min Scan# 904
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion: 49 Resp: 26650
Ion Ratio Lower Upper
49 100
129 2.6 0.0 4.6
130 81.0 69.0 103.4





#29
Tetrahydrofuran
Concen: 107.056 ug/l
RT: 7.350 min Scan# 906
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 42 Resp: 4155
Ion Ratio Lower Upper
42 100
72 49.0 40.7 61.1
71 46.6 37.4 56.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#30
Chloroform
Concen: 22.090 ug/l
RT: 7.508 min Scan# 933
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion: 83 Resp: 78446
Ion Ratio Lower Upper
83 100
85 57.4 52.1 78.1





#31
Cyclohexane
Concen: 21.128 ug/l
RT: 7.777 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion: 56 Resp: 6476
Ion Ratio Lower Upper
56 100
69 33.6 27.4 41.0
84 97.1 77.5 116.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#32
1,1,1-Trichloroethane
Concen: 21.214 ug/l
RT: 7.697 min Scan# 964
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17



Tgt Ion: 97 Resp: 68164
Ion Ratio Lower Upper
97 100
99 65.5 51.8 77.6
61 43.5 34.9 52.3





#33

1,2-Dichloroethane-d4

Concen: 48.684 ug/l

RT: 8.142 min Scan# 1037

Delta R.T. 0.006 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17

Instrument :

MSVOA_Y

ClientSampleId :

VY0513SBS01

Tgt Ion: 65 Resp: 8261

Ion Ratio Lower Upper





65 100

67 53.2 0.0 110.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17




Tgt Ion:114 Resp: 287281


Ion Ratio Lower Upper

114 100

63 19.3 0.0 36.4


88 14.4 0.0 27.4





#35
Dibromofluoromethane
Concen: 47.396 ug/l
RT: 7.716 min Scan# 966
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 113 Resp: 8172
Ion Ratio Lower Upper
113 100
111 102.7 81.8 122.6
192 17.5 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#36
1,1-Dichloropropene
Concen: 20.854 ug/l
RT: 7.911 min Scan# 999
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion: 75 Resp: 57492
Ion Ratio Lower Upper
75 100
110 35.7 18.1 54.1
77 31.9 24.3 36.5





#37
Ethyl Acetate
Concen: 21.870 ug/l
RT: 7.069 min Scan# 861
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 43 Resp: 3051
Ion Ratio Lower Upper
43 100
61 15.3 10.8 16.2
70 10.2 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#38
Carbon Tetrachloride
Concen: 20.910 ug/l
RT: 7.899 min Scan# 997
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Tgt Ion:117 Resp: 60831
Ion Ratio Lower Upper
117 100
119 91.8 78.5 117.7
121 30.5 25.8 38.8





#39

Methylcyclohexane

Concen: 21.003 ug/l

RT: 9.179 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17

Instrument :

MSVOA_Y

ClientSampleId :

VY0513SBS01





Tgt Ion: 83 Resp: 7361

Ion	Ratio	Lower	Upper
83	100		
55	65.5	53.4	80.0
98	48.9	41.0	61.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#40

Benzene

Concen: 20.689 ug/l

RT: 8.155 min Scan# 1039




Delta R.T. 0.000 min


Lab File: VY008744.D

Acq: 13 May 2022 13:17

Tgt Ion: 78 Resp: 165763


Ion	Ratio	Lower	Upper
78	100		
77	24.3	19.4	29.0





#41
Methacrylonitrile
Concen: 23.268 ug/l m
RT: 7.319 min Scan# 902
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 41 Resp: 15748
Ion Ratio Lower Upper
41 100
39 76.2 97.1 145.7
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#42
1,2-Dichloroethane
Concen: 22.183 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion: 62 Resp: 50295
Ion Ratio Lower Upper
62 100
98 10.4 0.0 20.6





#43
Isopropyl Acetate
Concen: 21.482 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 43 Resp: 5648
Ion Ratio Lower Upper
43 100
61 21.2 24.6 36.8
87 14.6 11.8 17.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#44
Trichloroethene
Concen: 21.365 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Tgt Ion:130 Resp: 45712
Ion Ratio Lower Upper
130 100
95 90.9 0.0 199.2





#45

1,2-Dichloropropane

Concen: 21.709 ug/l

RT: 9.215 min Scan# 1213

Delta R.T. 0.000 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17

Instrument :

MSVOA_Y

ClientSampleId :

VY0513SBS01

Tgt Ion: 63 Resp: 4199

Ion Ratio Lower Upper

63 100





65 29.9 24.7 37.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022

Supervised By :Semsettin Yesilyurt 05/16/2022



#46

Dibromomethane

Concen: 21.978 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. 0.000 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17




Tgt Ion: 93 Resp: 26978


Ion Ratio Lower Upper

93 100

95 84.5 67.4 101.2

174 89.9 74.2 111.2









#47
Bromodichloromethane
Concen: 21.695 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01

Tgt Ion: 83 Resp: 6171
Ion Ratio Lower Upper
83 100
85 66.4 52.2 78.4
127 8.9 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#48
Methyl methacrylate
Concen: 21.950 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Tgt Ion: 41 Resp: 24722
Ion Ratio Lower Upper
41 100
69 95.3 79.4 119.0
39 62.7 49.8 74.6





#49

1,4-Dioxane

Concen: 448.798 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. 0.000 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17

Instrument :

MSVOA_Y

ClientSampleId :

VY0513SBS01





Tgt Ion: 88 Resp: 690

Ion	Ratio	Lower	Upper
88	100		
43	27.9	24.2	36.2
58	61.8	45.7	68.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#50

Toluene-d8

Concen: 43.723 ug/l

RT: 10.179 min Scan# 1371




Delta R.T. 0.000 min


Lab File: VY008744.D

Acq: 13 May 2022 13:17

Tgt Ion: 98 Resp: 284305


Ion	Ratio	Lower	Upper
98	100		
100	65.6	52.8	79.2





#51
4-Methyl-2-Pentanone
Concen: 106.635 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion: 43 Resp: 150338
Ion Ratio Lower Upper
43 100
58 40.4 33.6 50.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#52
Toluene
Concen: 20.624 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion: 92 Resp: 109936
Ion Ratio Lower Upper
92 100
91 174.5 137.6 206.4





#53
t-1,3-Dichloropropene
Concen: 20.917 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion: 75 Resp: 61340
Ion Ratio Lower Upper
75 100
77 32.1 24.0 36.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#54
cis-1,3-Dichloropropene
Concen: 21.544 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion: 75 Resp: 70590
Ion Ratio Lower Upper
75 100
77 29.9 26.2 39.2
39 44.1 34.6 51.8





#55
1,1,2-Trichloroethane
Concen: 20.739 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01






Tgt Ion: 97 Resp: 34904


Ion	Ratio	Lower	Upper
97	100		
83	85.9	65.9	98.9
85	55.0	43.2	64.8
99	60.2	49.1	73.7

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022






#56
Ethyl methacrylate
Concen: 20.269 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17



Tgt Ion: 69 Resp: 43906

Ion	Ratio	Lower	Upper
69	100		
41	61.8	45.5	68.3
39	37.5	26.6	39.8





#57

1,3-Dichloropropane

Concen: 21.577 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17

Instrument :

MSVOA_Y

ClientSampleId :

VY0513SBS01

Tgt Ion: 76 Resp: 5957

Ion Ratio Lower Upper

76 100





78 33.3 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022

Supervised By :Semsettin Yesilyurt 05/16/2022



#58

2-Chloroethyl Vinyl ether

Concen: 98.694 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. 0.000 min

Lab File: VY008744.D




Acq: 13 May 2022 13:17


Tgt Ion: 63 Resp: 111961

Ion Ratio Lower Upper

63 100

106 30.9 23.5 35.3





#59

2-Hexanone

Concen: 106.445 ug/l

RT: 10.831 min Scan# 1478

Delta R.T. 0.000 min

Lab File: VY008744.D

Acq: 13 May 2022 13:17

Instrument :

MSVOA_Y

ClientSampleId :

VY0513SBS01

Tgt Ion: 43 Resp: 10113

Ion Ratio Lower Upper





43 100

58 56.1 29.6 88.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#60

Dibromochloromethane

Concen: 20.712 ug/l

RT: 10.983 min Scan# 1503

Delta R.T. 0.000 min

Lab File: VY008744.D




Acq: 13 May 2022 13:17


Tgt Ion:129 Resp: 41706

Ion Ratio Lower Upper

129 100


127 79.5 39.0 116.9





#61
1,2-Dibromoethane
Concen: 21.563 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:107 Resp: 3490
Ion Ratio Lower Upper
107 100
109 93.3 78.2 117.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#62
4-Bromofluorobenzene
Concen: 44.267 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion: 95 Resp: 99730
Ion Ratio Lower Upper
95 100
174 87.9 0.0 171.2
176 83.2 0.0 168.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:117 Resp: 271520
Ion Ratio Lower Upper
117 100
82 53.0 42.3 63.5
119 30.9 24.6 37.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#64
Tetrachloroethene
Concen: 20.239 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion:164 Resp: 41267
Ion Ratio Lower Upper
164 100
166 116.6 99.8 149.6
129 87.7 78.1 117.1
131 85.5 72.8 109.2





#65
Chlorobenzene
Concen: 20.285 ug/l
RT: 11.514 min Scan# 11590
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:112 Resp: 117145
Ion Ratio Lower Upper
112 100
114 29.7 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 20.020 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion:131 Resp: 42930
Ion Ratio Lower Upper
131 100
133 99.0 47.4 142.3
119 66.6 33.1 99.3





#67
Ethyl Benzene
Concen: 20.028 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion: 91 Resp: 20766
Ion Ratio Lower Upper
91 100
106 32.1 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#68
m/p-Xylenes
Concen: 39.419 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17



Tgt Ion:106 Resp: 162974
Ion Ratio Lower Upper
106 100
91 195.4 156.4 234.6









#69
o-Xylene
Concen: 19.873 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01

Tgt Ion:106 Resp: 77260
Ion Ratio Lower Upper
106 100
91 206.8 102.7 308.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#70
Styrene
Concen: 20.010 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion:104 Resp: 129508
Ion Ratio Lower Upper
104 100
78 46.2 35.9 53.9
103 53.9 43.1 64.7





#71
Bromoform
Concen: 20.701 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:173 Resp: 26125
Ion Ratio Lower Upper
173 100
175 45.8 24.0 72.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion:152 Resp: 129997
Ion Ratio Lower Upper
152 100
115 58.5 28.2 84.7
150 165.7 0.0 348.2





#73
Isopropylbenzene
Concen: 20.264 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:105 Resp: 199279
Ion Ratio Lower Upper
105 100
120 26.8 13.4 40.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#74
N-ethyl acetate
Concen: 21.444 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion: 43 Resp: 48751
Ion Ratio Lower Upper
43 100
70 47.6 39.0 58.4
55 26.8 20.2 30.4
61 27.0 21.0 31.4





#75
1,1,2,2-Tetrachloroethane
Concen: 21.189 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion: 83 Resp: 4280
Ion Ratio Lower Upper
83 100
131 9.6 5.0 14.9
85 64.3 32.2 96.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#76
1,2,3-Trichloropropane
Concen: 19.228 ug/l m
RT: 12.636 min Scan# 1774
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17



Tgt Ion: 75 Resp: 25153
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 20.550 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01

Tgt Ion:156 Resp: 45758
Ion Ratio Lower Upper
156 100
77 186.1 91.1 273.3
158 96.4 48.2 144.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#78
n-propylbenzene
Concen: 20.820 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Tgt Ion: 91 Resp: 247981
Ion Ratio Lower Upper
91 100
120 22.1 11.6 34.8









#79
2-Chlorotoluene
Concen: 20.791 ug/l
RT: 12.758 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01

Tgt Ion: 91 Resp: 13333
Ion Ratio Lower Upper
91 100
126 35.3 17.8 53.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#80
1,3,5-Trimethylbenzene
Concen: 20.623 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion:105 Resp: 165374
Ion Ratio Lower Upper
105 100
120 50.4 25.1 75.4





#81
trans-1,4-Dichloro-2-butene
Concen: 21.825 ug/l
RT: 12.380 min Scan# 1732
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01



Tgt Ion: 75 Resp: 14778
Ion Ratio Lower Upper
75 100
53 85.8 66.3 99.5
89 45.0 36.8 55.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#82
4-Chlorotoluene
Concen: 20.741 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Tgt Ion: 91 Resp: 140038
Ion Ratio Lower Upper
91 100
126 36.1 17.9 53.8









#83
tert-Butylbenzene
Concen: 20.207 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01

Tgt Ion:119 Resp: 143964
Ion Ratio Lower Upper
119 100
91 62.4 31.2 93.6
134 27.4 13.5 40.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#84
1,2,4-Trimethylbenzene
Concen: 20.733 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion:105 Resp: 165239
Ion Ratio Lower Upper
105 100
120 46.2 23.2 69.6





#85
 sec-Butylbenzene
 Concen: 20.498 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VY008744.D
 Acq: 13 May 2022 13:17




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0513SBS01




Tgt Ion:105 Resp: 221586
 Ion Ratio Lower Upper
 105 100
 134 20.4 10.1 30.1

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
 Supervised By :Semsettin Yesilyurt 05/16/2022




#86
 p-Isopropyltoluene
 Concen: 20.347 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY008744.D
 Acq: 13 May 2022 13:17




Tgt Ion:119 Resp: 183582
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.3 39.9
 91 22.8 11.1 33.3





#87
1,3-Dichlorobenzene
Concen: 20.337 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:146 Resp: 94075
Ion Ratio Lower Upper
146 100
111 39.1 19.3 57.9
148 64.7 31.4 94.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#88
1,4-Dichlorobenzene
Concen: 20.987 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17



Tgt Ion:146 Resp: 95282
Ion Ratio Lower Upper
146 100
111 37.8 19.0 57.0
148 62.6 32.0 96.0









#89
n-Butylbenzene
Concen: 20.798 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17

Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01

Tgt Ion: 91 Resp: 17551
Ion Ratio Lower Upper
91 100
92 52.3 26.4 79.2
134 24.8 13.2 39.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#90
Hexachloroethane
Concen: 20.922 ug/l
RT: 13.959 min Scan# 1991
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17


Tgt Ion:117 Resp: 34825
Ion Ratio Lower Upper
117 100
201 66.9 34.4 103.1





#91
1,2-Dichlorobenzene
Concen: 20.540 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:146 Resp: 82290
Ion Ratio Lower Upper
146 100
111 39.5 20.3 60.8
148 63.8 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 20.977 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion: 75 Resp: 6291
Ion Ratio Lower Upper
75 100
155 85.5 44.7 134.1
157 105.5 60.5 181.4





#93
1,2,4-Trichlorobenzene
Concen: 20.991 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:180 Resp: 4670
Ion Ratio Lower Upper
180 100
182 94.9 46.8 140.4
145 31.2 15.0 45.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022




#94
Hexachlorobutadiene
Concen: 20.845 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Tgt Ion:225 Resp: 24417
Ion Ratio Lower Upper
225 100
223 64.6 31.5 94.5
227 60.3 31.2 93.6





#95
Naphthalene
Concen: 20.594 ug/l
RT: 15.239 min Scan# 21
Delta R.T. 0.000 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17




Instrument :
MSVOA_Y
ClientSampleId :
VY0513SBS01




Tgt Ion:128 Resp: 9793
Ion Ratio Lower Upper
128 100
127 13.2 10.2 15.2
129 11.4 8.9 13.3

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/16/2022
Supervised By :Semsettin Yesilyurt 05/16/2022



#96
1,2,3-Trichlorobenzene
Concen: 20.630 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.006 min
Lab File: VY008744.D
Acq: 13 May 2022 13:17



Tgt Ion:180 Resp: 40640
Ion Ratio Lower Upper
180 100
182 95.4 46.9 140.7
145 32.5 15.6 46.8

