

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
 Data File : VY013741.D
 Acq On : 15 May 2023 17:14
 Operator : KP/MD
 Sample : 02750-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PAT-1-051123

Quant Time: May 16 01:33:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	172130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	298177	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	290714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	131279	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	114202	60.523	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 121.040%	
35) Dibromofluoromethane	7.716	113	103289	55.076	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.160%	
50) Toluene-d8	10.179	98	370288	55.502	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 111.000%	
62) 4-Bromofluorobenzene	12.477	95	126224	56.121	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 112.240%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	34162	7.856	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
Data File : VY013741.D
Acq On : 15 May 2023 17:14
Operator : KP/MD
Sample : 02750-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PAT-1-051123

Quant Time: May 16 01:33:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

