

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
 Data File : VY013742.D
 Acq On : 15 May 2023 17:37
 Operator : KP/MD
 Sample : 02750-03
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PAT-2-051123

Quant Time: May 16 01:33:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	177374	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	313664	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	303278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	132544	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	117293	60.323	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.640%	
35) Dibromofluoromethane	7.715	113	106951	54.213	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.420%	
50) Toluene-d8	10.178	98	388814	55.402	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 110.800%	
62) 4-Bromofluorobenzene	12.477	95	125523	53.054	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 106.100%	
Target Compounds						
20) Methylene Chloride	4.716	84	33419	6.969	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
Data File : VY013742.D
Acq On : 15 May 2023 17:37
Operator : KP/MD
Sample : 02750-03
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PAT-2-051123

Quant Time: May 16 01:33:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

