

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
 Data File : VY013744.D
 Acq On : 15 May 2023 18:24
 Operator : KP/MD
 Sample : 02760-03
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-4-051123

Quant Time: May 16 01:34:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	180979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	317415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	307591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	145229	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	121226	61.104	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.200%
35) Dibromofluoromethane	7.716	113	106977	53.585	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.180%
50) Toluene-d8	10.179	98	388768	54.741	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.480%
62) 4-Bromofluorobenzene	12.477	95	130902	54.673	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
Data File : VY013744.D
Acq On : 15 May 2023 18:24
Operator : KP/MD
Sample : 02760-03
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-4-051123

Quant Time: May 16 01:34:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

