

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
 Data File : VY013747.D
 Acq On : 15 May 2023 19:33
 Operator : KP/MD
 Sample : 02723-04MSD
 Misc : 5.65g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-14-G1-SO-14-050923MSD

Quant Time: May 16 01:35:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 05/16/2023
 Supervised By :Mahesh Dadoda 05/17/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	148738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	230256	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	207053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	108630	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	92045	56.452	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 112.900%			
35) Dibromofluoromethane	7.716	113	81659	56.386	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 112.780%			
50) Toluene-d8	10.179	98	269063	52.226	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 104.460%			
62) 4-Bromofluorobenzene	12.477	95	101193	58.263	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 116.520%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.906	85	75197	53.179	ug/l	99
3) Chloromethane	2.113	50	102178	53.908	ug/l	97
4) Vinyl Chloride	2.253	62	106695	56.535	ug/l	97
5) Bromomethane	2.650	94	79575	52.912	ug/l	97
6) Chloroethane	2.796	64	67535	51.800	ug/l	97
7) Trichlorofluoromethane	3.125	101	159767	55.883	ug/l	95
8) Diethyl Ether	3.527	74	53106	54.547	ug/l	84
9) 1,1,2-Trichlorotrifluo...	3.905	101	99516	62.056	ug/l	95
10) Methyl Iodide	4.094	142	93152	46.248	ug/l	99
11) Tert butyl alcohol	4.978	59	50131	272.912	ug/l #	90
12) 1,1-Dichloroethene	3.875	96	83759	54.185	ug/l	92
13) Acrolein	3.723	56	12269	80.103	ug/l	98
14) Allyl chloride	4.478	41	158997	55.433	ug/l	96
15) Acrylonitrile	5.167	53	173710	318.593	ug/l	98
16) Acetone	3.954	43	164372	255.989	ug/l	100
17) Carbon Disulfide	4.192	76	232727	46.348	ug/l	99
18) Methyl Acetate	4.478	43	154668	72.931	ug/l	94
19) Methyl tert-butyl Ether	5.216	73	261711	58.644	ug/l	99
20) Methylene Chloride	4.716	84	108728	54.834	ug/l #	84
21) trans-1,2-Dichloroethene	5.222	96	90432	53.505	ug/l	96
22) Diisopropyl ether	6.125	45	337699	60.237	ug/l #	92
23) Vinyl Acetate	6.064	43	908868	286.328	ug/l	95
24) 1,1-Dichloroethane	6.021	63	182849	58.520	ug/l	98
25) 2-Butanone	6.990	43	232813	285.895	ug/l	90
26) 2,2-Dichloropropane	6.978	77	152534	56.654	ug/l	100
27) cis-1,2-Dichloroethene	6.984	96	128000	68.544	ug/l	95
28) Bromochloromethane	7.338	49	64281	57.975	ug/l #	79
29) Tetrahydrofuran	7.350	42	151842	299.548	ug/l	91
30) Chloroform	7.508	83	181886	58.549	ug/l	100
31) Cyclohexane	7.789	56	152026	52.077	ug/l	90
32) 1,1,1-Trichloroethane	7.704	97	158504	57.177	ug/l	97
36) 1,1-Dichloropropene	7.917	75	134683	56.036	ug/l	97
37) Ethyl Acetate	7.076	43	105342	58.241	ug/l	96
38) Carbon Tetrachloride	7.899	117	145359	57.702	ug/l	99
39) Methylcyclohexane	9.185	83	156014	55.061	ug/l	95
40) Benzene	8.161	78	379306	56.024	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
 Data File : VY013747.D
 Acq On : 15 May 2023 19:33
 Operator : KP/MD
 Sample : 02723-04MSD
 Misc : 5.65g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-14-G1-SO-14-050923MSD

Quant Time: May 16 01:35:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 05/16/2023
 Supervised By :Mahesh Dadoda 05/17/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.313	41	51618m	61.242	ug/l	
42) 1,2-Dichloroethane	8.234	62	133166	60.162	ug/l	98
43) Isopropyl Acetate	8.271	43	185290	60.693	ug/l	95
44) Trichloroethene	8.941	130	111392	61.853	ug/l	92
45) 1,2-Dichloropropane	9.215	63	105450	59.569	ug/l	97
46) Dibromomethane	9.307	93	62894	59.380	ug/l	93
47) Bromodichloromethane	9.496	83	142435	60.098	ug/l	99
48) Methyl methacrylate	9.289	41	90754	64.727	ug/l	90
49) 1,4-Dioxane	9.301	88	16157	1068.064	ug/l #	87
51) 4-Methyl-2-Pentanone	10.069	43	538510	316.542	ug/l	95
52) Toluene	10.240	92	235931	57.908	ug/l	100
53) t-1,3-Dichloropropene	10.465	75	148619	59.250	ug/l	98
54) cis-1,3-Dichloropropene	9.929	75	163280	57.850	ug/l	95
55) 1,1,2-Trichloroethane	10.642	97	86213	61.463	ug/l	97
56) Ethyl methacrylate	10.508	69	117039	60.144	ug/l	91
57) 1,3-Dichloropropane	10.788	76	145744	61.352	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.782	63	204546	280.309	ug/l	92
59) 2-Hexanone	10.831	43	371794	300.645	ug/l	96
60) Dibromochloromethane	10.983	129	105060	61.685	ug/l	100
61) 1,2-Dibromoethane	11.087	107	83750	60.730	ug/l	100
64) Tetrachloroethene	10.721	164	82345	52.301	ug/l	93
65) Chlorobenzene	11.514	112	250070	55.443	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.587	131	98457	58.367	ug/l	100
67) Ethyl Benzene	11.593	91	449203	56.716	ug/l	98
68) m/p-Xylenes	11.703	106	339118	112.629	ug/l	96
69) o-Xylene	12.026	106	161805	56.822	ug/l	96
70) Styrene	12.044	104	279076	58.614	ug/l	100
71) Bromoform	12.203	173	72238	61.137	ug/l #	99
73) Isopropylbenzene	12.325	105	442511	53.857	ug/l	98
74) N-amyl acetate	12.142	43	154298	52.772	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.581	83	112783	56.158	ug/l	99
76) 1,2,3-Trichloropropane	12.629	75	82679m	56.436	ug/l	
77) Bromobenzene	12.605	156	106347	53.418	ug/l	91
78) n-propylbenzene	12.672	91	541604	53.516	ug/l	97
79) 2-Chlorotoluene	12.757	91	303848	53.764	ug/l	98
80) 1,3,5-Trimethylbenzene	12.812	105	362593	54.069	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.373	75	38381	53.591	ug/l	95
82) 4-Chlorotoluene	12.855	91	310277	52.674	ug/l	97
83) tert-Butylbenzene	13.075	119	323443	55.670	ug/l	99
84) 1,2,4-Trimethylbenzene	13.117	105	356574	54.689	ug/l	100
85) sec-Butylbenzene	13.251	105	481270	55.193	ug/l	98
86) p-Isopropyltoluene	13.367	119	388065	55.341	ug/l	100
87) 1,3-Dichlorobenzene	13.367	146	201693	53.176	ug/l	100
88) 1,4-Dichlorobenzene	13.446	146	200515	52.334	ug/l	100
89) n-Butylbenzene	13.696	91	370150	54.084	ug/l	98
90) Hexachloroethane	13.959	117	78345	53.086	ug/l	98
91) 1,2-Dichlorobenzene	13.739	146	185174	54.063	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	19347	54.546	ug/l	91
93) 1,2,4-Trichlorobenzene	15.007	180	108656	53.208	ug/l	99
94) Hexachlorobutadiene	15.111	225	65950	55.492	ug/l	100
95) Naphthalene	15.233	128	247500	55.860	ug/l	100
96) 1,2,3-Trichlorobenzene	15.422	180	100101	54.998	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
Data File : VY013747.D
Acq On : 15 May 2023 19:33
Operator : KP/MD
Sample : 02723-04MSD
Misc : 5.65g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023

Quant Time: May 16 01:35:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051523\
 Data File : VY013747.D
 Acq On : 15 May 2023 19:33
 Operator : KP/MD
 Sample : 02723-04MSD
 Misc : 5.65g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-14-G1-SO-14-050923MSD

Quant Time: May 16 01:35:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Krupa Patel 05/16/2023
 Supervised By :Mahesh Dadoda 05/17/2023





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD



Tgt Ion:168 Resp: 148738
Ion Ratio Lower Upper
168 100
99 53.6 41.8 62.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#2
Dichlorodifluoromethane
Concen: 53.179 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 85 Resp: 75197
Ion Ratio Lower Upper
85 100
87 31.4 16.1 48.2





#3
Chloromethane
Concen: 53.908 ug/l
RT: 2.113 min Scan# 48
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion: 50 Resp: 102178
Ion Ratio Lower Upper
50 100
52 30.5 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#4
Vinyl Chloride
Concen: 56.535 ug/l
RT: 2.253 min Scan# 71
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Tgt Ion: 62 Resp: 106695
Ion Ratio Lower Upper
62 100
64 30.1 25.4 38.0





#5
Bromomethane
Concen: 52.912 ug/l
RT: 2.650 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD



Tgt Ion: 94 Resp: 7957
Ion Ratio Lower Upper
94 100
96 89.8 74.2 111.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#6
Chloroethane
Concen: 51.800 ug/l
RT: 2.796 min Scan# 160
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 64 Resp: 67535
Ion Ratio Lower Upper
64 100
66 29.4 24.6 37.0





#7
Trichlorofluoromethane
Concen: 55.883 ug/l
RT: 3.125 min Scan# 214
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion:101 Resp: 15976
Ion Ratio Lower Upper
101 100
103 68.6 51.6 77.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#8
Diethyl Ether
Concen: 54.547 ug/l
RT: 3.527 min Scan# 280
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33



Tgt Ion: 74 Resp: 53106
Ion Ratio Lower Upper
74 100
45 111.7 48.2 144.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 62.056 ug/l
RT: 3.905 min Scan# 341
Delta R.T. 0.012 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33

Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD

Tgt Ion:101 Resp: 99510
Ion Ratio Lower Upper
101 100
85 46.9 35.6 53.4
151 79.2 68.0 102.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#10
Methyl Iodide
Concen: 46.248 ug/l
RT: 4.094 min Scan# 373
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion:142 Resp: 93152
Ion Ratio Lower Upper
142 100
127 45.5 36.1 54.1
141 14.9 11.6 17.4





#11
Tert butyl alcohol
Concen: 272.912 ug/l
RT: 4.978 min Scan# 519
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion: 59 Resp: 5013
Ion Ratio Lower Upper
59 100
57 7.1 3.0 4.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#12
1,1-Dichloroethene
Concen: 54.185 ug/l
RT: 3.875 min Scan# 337
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33



Tgt Ion: 96 Resp: 83759
Ion Ratio Lower Upper
96 100
61 168.4 123.9 185.9
98 59.4 48.9 73.3





#13

Acrolein

Concen: 80.103 ug/l

RT: 3.723 min Scan# 314

Delta R.T. -0.006 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD

Tgt Ion: 56 Resp: 12269

Ion Ratio Lower Upper

56 100





55 69.3 54.1 81.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#14

Allyl chloride

Concen: 55.433 ug/l

RT: 4.478 min Scan# 436

Delta R.T. 0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33




Tgt Ion: 41 Resp: 158997


Ion Ratio Lower Upper

41 100

39 70.1 56.6 85.0


76 32.6 30.5 45.7





#15
Acrylonitrile
Concen: 318.593 ug/l
RT: 5.167 min Scan# 54
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion: 53 Resp: 173710
Ion Ratio Lower Upper
53 100
52 83.9 65.4 98.0
51 37.9 29.3 43.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#16
Acetone
Concen: 255.989 ug/l
RT: 3.954 min Scan# 350
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33



Tgt Ion: 43 Resp: 164372
Ion Ratio Lower Upper
43 100
58 31.6 25.4 38.0









#17
Carbon Disulfide
Concen: 46.348 ug/l
RT: 4.192 min Scan# 389
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33

Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD

Tgt Ion: 76 Resp: 23272
Ion Ratio Lower Upper
76 100
78 8.6 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#18
Methyl Acetate
Concen: 72.931 ug/l
RT: 4.478 min Scan# 436
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33

Tgt Ion: 43 Resp: 154668
Ion Ratio Lower Upper
43 100
74 21.1 19.4 29.2





#19

Methyl tert-butyl Ether

Concen: 58.644 ug/l

RT: 5.216 min Scan# 51

Delta R.T. 0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD

Tgt Ion: 73 Resp: 26171

Ion Ratio Lower Upper

73 100





57 22.7 17.8 26.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#20

Methylene Chloride

Concen: 54.834 ug/l

RT: 4.716 min Scan# 475

Delta R.T. 0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Tgt Ion: 84 Resp: 108728




Ion Ratio Lower Upper


84 100

49 136.1 88.0 132.0#

51 42.0 27.4 41.0#


86 61.9 50.4 75.6





#21
trans-1,2-Dichloroethene
Concen: 53.505 ug/l
RT: 5.222 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion: 96 Resp: 9043
Ion Ratio Lower Upper
96 100
61 144.2 110.6 165.8
98 67.8 53.1 79.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#22
Diisopropyl ether
Concen: 60.237 ug/l
RT: 6.125 min Scan# 706
Delta R.T. 0.007 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Tgt Ion: 45 Resp: 337699
Ion Ratio Lower Upper
45 100
43 56.2 48.4 72.6
87 22.2 23.8 35.6#
59 11.0 9.6 14.4





#23
 Vinyl Acetate
 Concen: 286.328 ug/l
 RT: 6.064 min Scan# 695
 Delta R.T. 0.007 min
 Lab File: VY013747.D
 Acq: 15 May 2023 19:33




Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-14-G1-SO-14-050923MSD




Tgt Ion: 43 Resp: 90886
 Ion Ratio Lower Upper
 43 100
 86 8.7 8.5 12.7

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 05/16/2023
 Supervised By :Mahesh Dadoda 05/17/2023




#24
 1,1-Dichloroethane
 Concen: 58.520 ug/l
 RT: 6.021 min Scan# 689
 Delta R.T. 0.006 min
 Lab File: VY013747.D
 Acq: 15 May 2023 19:33




Tgt Ion: 63 Resp: 182849
 Ion Ratio Lower Upper
 63 100
 98 6.1 3.5 10.6
 100 3.9 2.2 6.6





#25
2-Butanone
Concen: 285.895 ug/l
RT: 6.990 min Scan# 848
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD



Tgt Ion: 43 Resp: 23281
Ion Ratio Lower Upper
43 100
72 22.3 21.9 32.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#26
2,2-Dichloropropane
Concen: 56.654 ug/l
RT: 6.978 min Scan# 846
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 77 Resp: 152534
Ion Ratio Lower Upper
77 100
97 23.7 11.8 35.4





#27
 cis-1,2-Dichloroethene
 Concen: 68.544 ug/l
 RT: 6.984 min Scan# 847
 Delta R.T. 0.000 min
 Lab File: VY013747.D
 Acq: 15 May 2023 19:33




Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-14-G1-SO-14-050923MSD



Tgt Ion: 96 Resp: 128000
 Ion Ratio Lower Upper
 96 100
 61 150.7 0.0 285.0
 98 65.4 0.0 130.6




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 05/16/2023
 Supervised By :Mahesh Dadoda 05/17/2023



#28
 Bromochloromethane
 Concen: 57.975 ug/l
 RT: 7.338 min Scan# 905
 Delta R.T. 0.006 min
 Lab File: VY013747.D
 Acq: 15 May 2023 19:33

Tgt Ion: 49 Resp: 64281
 Ion Ratio Lower Upper
 49 100
 129 2.1 0.0 5.2
 130 69.6 72.1 108.1#





#29

Tetrahydrofuran

Concen: 299.548 ug/l

RT: 7.350 min Scan# 906

Delta R.T. 0.006 min

Lab File: VY013747.D


Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD






Tgt Ion:	42	Resp:	15184
Ion Ratio	100	Lower	Upper
72	39.7	35.9	53.9
71	36.8	34.3	51.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#30

Chloroform

Concen: 58.549 ug/l




RT: 7.508 min Scan# 933


Delta R.T. 0.006 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33


Tgt Ion:	83	Resp:	181886
Ion Ratio	100	Lower	Upper
85	64.7	52.0	78.0





#31
Cyclohexane
Concen: 52.077 ug/l
RT: 7.789 min Scan# 917
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion: 56 Resp: 152020
Ion Ratio Lower Upper
56 100
69 33.0 29.2 43.8
84 80.7 73.0 109.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#32
1,1,1-Trichloroethane
Concen: 57.177 ug/l
RT: 7.704 min Scan# 965
Delta R.T. 0.007 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33



Tgt Ion: 97 Resp: 158504
Ion Ratio Lower Upper
97 100
99 65.8 52.2 78.2
61 46.2 33.8 50.8





#33

1,2-Dichloroethane-d4

Concen: 56.452 ug/l

RT: 8.142 min Scan# 1037

Delta R.T. 0.006 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD

Tgt Ion: 65 Resp: 92045

Ion Ratio Lower Upper

65 100





67 53.4 0.0 103.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33




Tgt Ion:114 Resp: 230256


Ion Ratio Lower Upper

114 100

63 21.1 0.0 36.6


88 14.4 0.0 28.8





#35
Dibromofluoromethane
Concen: 56.386 ug/l
RT: 7.716 min Scan# 90
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD



Tgt Ion: 113 Resp: 81659
Ion Ratio Lower Upper
113 100
111 103.1 82.2 123.2
192 20.4 16.7 25.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#36
1,1-Dichloropropene
Concen: 56.036 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 75 Resp: 134683
Ion Ratio Lower Upper
75 100
110 34.4 18.7 56.1
77 31.1 24.6 37.0





#37
Ethyl Acetate
Concen: 58.241 ug/l
RT: 7.076 min Scan# 862
Delta R.T. 0.007 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD



Tgt Ion: 43 Resp: 105342
Ion Ratio Lower Upper
43 100
61 12.6 11.4 17.0
70 8.4 7.9 11.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#38
Carbon Tetrachloride
Concen: 57.702 ug/l
RT: 7.899 min Scan# 997
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion:117 Resp: 145359
Ion Ratio Lower Upper
117 100
119 98.5 79.0 118.6
121 30.4 26.6 40.0





#39
Methylcyclohexane
Concen: 55.061 ug/l
RT: 9.185 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD



Tgt Ion: 83 Resp: 156014
Ion Ratio Lower Upper
83 100
55 80.3 59.0 88.4
98 45.5 37.0 55.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#40
Benzene
Concen: 56.024 ug/l
RT: 8.161 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 78 Resp: 379306
Ion Ratio Lower Upper
78 100
77 24.1 19.2 28.8





#41
Methacrylonitrile
Concen: 61.242 ug/l m
RT: 7.313 min Scan# 901
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion: 41 Resp: 51618
Ion Ratio Lower Upper
41 100
39 0.0 44.2 66.4
67 0.0 62.9 94.3
52 0.0 26.4 39.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#42
1,2-Dichloroethane
Concen: 60.162 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Tgt Ion: 62 Resp: 133166
Ion Ratio Lower Upper
62 100
98 8.3 0.0 17.8





#43
Isopropyl Acetate
Concen: 60.693 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. 0.001 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD



Tgt Ion: 43 Resp: 185290
Ion Ratio Lower Upper
43 100
61 25.2 22.2 33.4
87 11.5 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#44
Trichloroethene
Concen: 61.853 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33

Tgt Ion:130 Resp: 111392
Ion Ratio Lower Upper
130 100
95 102.4 0.0 188.8





#45

1,2-Dichloropropane

Concen: 59.569 ug/l

RT: 9.215 min Scan# 1213

Delta R.T. 0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD

Tgt Ion: 63 Resp: 105450

Ion Ratio Lower Upper

63 100





65 29.2 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#46

Dibromomethane

Concen: 59.380 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33




Tgt Ion: 93 Resp: 62894


Ion Ratio Lower Upper

93 100

95 80.6 67.1 100.7


174 98.2 87.0 130.4





#47
Bromodichloromethane
Concen: 60.098 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD



Tgt Ion: 83 Resp: 142435
Ion Ratio Lower Upper
83 100
85 64.8 51.4 77.2
127 8.6 8.1 12.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#48
Methyl methacrylate
Concen: 64.727 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33

Tgt Ion: 41 Resp: 90754
Ion Ratio Lower Upper
41 100
69 76.4 70.2 105.4
39 55.3 47.9 71.9





#49

1,4-Dioxane

Concen: 1068.064 ug/l

RT: 9.301 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD





Tgt Ion:	88	Resp:	1615
Ion Ratio	Lower	Upper	
88	100		
43	33.2	16.8	25.2
58	75.6	54.3	81.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#50

Toluene-d8

Concen: 52.226 ug/l




RT: 10.179 min Scan# 1371


Delta R.T. -0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33


Tgt Ion:	98	Resp:	269063
Ion Ratio	Lower	Upper	
98	100		
100	62.2	50.5	75.7





#51
4-Methyl-2-Pentanone
Concen: 316.542 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD




Tgt Ion: 43 Resp: 538510
Ion Ratio Lower Upper
43 100
58 35.2 30.6 46.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#52
Toluene
Concen: 57.908 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Tgt Ion: 92 Resp: 235931
Ion Ratio Lower Upper
92 100
91 175.7 140.2 210.4





#53
t-1,3-Dichloropropene
Concen: 59.250 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD



Tgt Ion: 75 Resp: 148619
Ion Ratio Lower Upper
75 100
77 31.4 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#54
cis-1,3-Dichloropropene
Concen: 57.850 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 75 Resp: 163280
Ion Ratio Lower Upper
75 100
77 31.8 24.8 37.2
39 52.9 38.1 57.1





#55
1,1,2-Trichloroethane
Concen: 61.463 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD



Tgt Ion	Ratio	Lower	Upper
97	100		
83	86.4	66.4	99.6
85	54.7	45.6	68.4
99	62.3	48.6	72.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#56
Ethyl methacrylate
Concen: 60.144 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion	Ratio	Lower	Upper
69	100		
41	73.2	51.4	77.0
39	39.6	29.0	43.6





#57
1,3-Dichloropropane
Concen: 61.352 ug/l
RT: 10.788 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD



Tgt Ion: 76 Resp: 145744
Ion Ratio Lower Upper
76 100
78 32.1 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#58
2-Chloroethyl Vinyl ether
Concen: 280.309 ug/l
RT: 9.782 min Scan# 1306
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 63 Resp: 204546
Ion Ratio Lower Upper
63 100
106 25.0 23.5 35.3





#59
2-Hexanone
Concen: 300.645 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD



Tgt Ion: 43 Resp: 371794
Ion Ratio Lower Upper
43 100
58 49.4 26.3 78.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#60
Dibromochloromethane
Concen: 61.685 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion:129 Resp: 105060
Ion Ratio Lower Upper
129 100
127 77.5 38.7 116.1





#61
1,2-Dibromoethane
Concen: 60.730 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion:107 Resp: 83750
Ion Ratio Lower Upper
107 100
109 94.0 75.6 113.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#62
4-Bromofluorobenzene
Concen: 58.263 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Tgt Ion: 95 Resp: 101193
Ion Ratio Lower Upper
95 100
174 87.0 0.0 182.0
176 83.9 0.0 175.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11586
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD



Tgt Ion:117 Resp: 20705
Ion Ratio Lower Upper
117 100
82 57.0 43.0 64.6
119 32.1 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#64
Tetrachloroethene
Concen: 52.301 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion:164 Resp: 82345
Ion Ratio Lower Upper
164 100
166 134.3 97.5 146.3
129 95.7 72.6 109.0
131 90.2 69.8 104.6





#65
Chlorobenzene
Concen: 55.443 ug/l
RT: 11.514 min Scan# 11590
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD




Tgt Ion:112 Resp: 250070
Ion Ratio Lower Upper
112 100
114 31.9 24.9 37.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#66
1,1,1,2-Tetrachloroethane
Concen: 58.367 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33



Tgt Ion:131 Resp: 98457
Ion Ratio Lower Upper
131 100
133 95.6 47.7 143.1
119 65.2 32.6 98.0





#67

Ethyl Benzene

Concen: 56.716 ug/l

RT: 11.593 min Scan# 1603

Delta R.T. 0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD

Tgt Ion: 91 Resp: 44920

Ion Ratio Lower Upper

91 100





106 30.3 25.1 37.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#68

m/p-Xylenes

Concen: 112.629 ug/l

RT: 11.703 min Scan# 1621

Delta R.T. -0.000 min

Lab File: VY013747.D




Acq: 15 May 2023 19:33


Tgt Ion:106 Resp: 339118

Ion Ratio Lower Upper

106 100

91 207.0 160.9 241.3









#69
o-Xylene
Concen: 56.822 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33

Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD

Tgt Ion:106 Resp: 161809
Ion Ratio Lower Upper
106 100
91 217.5 105.5 316.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#70
Styrene
Concen: 58.614 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33

Tgt Ion:104 Resp: 279076
Ion Ratio Lower Upper
104 100
78 50.5 40.3 60.5
103 55.7 45.0 67.4





#71

Bromoform

Concen: 61.137 ug/l

RT: 12.203 min Scan# 1703

Delta R.T. -0.006 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD





Tgt Ion:173	Resp:	72238
Ion Ratio	Lower	Upper
173	100	
175	49.1	24.3 72.8
254	0.1	0.2 0.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.422 min Scan# 1903


Delta R.T. 0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33


Tgt Ion:152	Resp:	108630
Ion Ratio	Lower	Upper
152	100	
115	58.5	28.1 84.2
150	171.6	0.0 352.0





#73
Isopropylbenzene
Concen: 53.857 ug/l
RT: 12.325 min Scan# 1723
Delta R.T. -0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD



Tgt Ion:105 Resp: 44251
Ion Ratio Lower Upper
105 100
120 26.4 13.7 41.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#74
N-ethyl acetate
Concen: 52.772 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 43 Resp: 154298
Ion Ratio Lower Upper
43 100
70 36.5 33.0 49.4
55 23.2 19.8 29.6
61 21.7 19.1 28.7





#75
1,1,2,2-Tetrachloroethane
Concen: 56.158 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD



Tgt Ion: 83 Resp: 11278
Ion Ratio Lower Upper
83 100
131 10.8 5.7 17.0
85 64.2 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#76
1,2,3-Trichloropropane
Concen: 56.436 ug/l m
RT: 12.629 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 75 Resp: 82679
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 53.418 ug/l
RT: 12.605 min Scan# 1769
Delta R.T. -0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion:156 Resp: 10634
Ion Ratio Lower Upper
156 100
77 194.6 87.6 262.9
158 96.9 48.5 145.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#78
n-propylbenzene
Concen: 53.516 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Tgt Ion: 91 Resp: 541604
Ion Ratio Lower Upper
91 100
120 22.3 11.9 35.9





#79
2-Chlorotoluene
Concen: 53.764 ug/l
RT: 12.757 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion: 91 Resp: 303848
Ion Ratio Lower Upper
91 100
126 34.6 17.8 53.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#80
1,3,5-Trimethylbenzene
Concen: 54.069 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33



Tgt Ion:105 Resp: 362593
Ion Ratio Lower Upper
105 100
120 48.1 24.7 74.1









#81
trans-1,4-Dichloro-2-butene
Concen: 53.591 ug/l
RT: 12.373 min Scan# 1731
Delta R.T. -0.006 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33

Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD

Tgt Ion: 75 Resp: 3838
Ion Ratio Lower Upper
75 100
53 96.1 73.2 109.8
89 43.9 38.0 57.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023



#82
4-Chlorotoluene
Concen: 52.674 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33


Tgt Ion: 91 Resp: 310277
Ion Ratio Lower Upper
91 100
126 34.4 17.9 53.8





#83
tert-Butylbenzene
Concen: 55.670 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. 0.001 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument :
MSVOA_Y
ClientSampleId :
CB-14-G1-SO-14-050923MSD




Tgt Ion:119 Resp: 32344
Ion Ratio Lower Upper
119 100
91 64.9 32.2 96.6
134 25.4 13.0 38.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#84
1,2,4-Trimethylbenzene
Concen: 54.689 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Tgt Ion:105 Resp: 356574
Ion Ratio Lower Upper
105 100
120 44.6 22.3 66.9





#85
 sec-Butylbenzene
 Concen: 55.193 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VY013747.D
 Acq: 15 May 2023 19:33




Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-14-G1-SO-14-050923MSD




Tgt Ion:105 Resp: 481270
 Ion Ratio Lower Upper
 105 100
 134 19.8 10.3 30.9

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 05/16/2023
 Supervised By :Mahesh Dadoda 05/17/2023




#86
 p-Isopropyltoluene
 Concen: 55.341 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY013747.D
 Acq: 15 May 2023 19:33




Tgt Ion:119 Resp: 388065
 Ion Ratio Lower Upper
 119 100
 134 26.6 13.4 40.2
 91 24.1 12.1 36.3





#87
1,3-Dichlorobenzene
Concen: 53.176 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD




Tgt Ion:146 Resp: 20169
Ion Ratio Lower Upper
146 100
111 38.5 19.3 57.9
148 63.9 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#88
1,4-Dichlorobenzene
Concen: 52.334 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33



Tgt Ion:146 Resp: 200515
Ion Ratio Lower Upper
146 100
111 37.8 18.9 56.5
148 63.8 31.9 95.5





#89

n-Butylbenzene

Concen: 54.084 ug/l

RT: 13.696 min Scan# 1948

Delta R.T. 0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD





Tgt Ion: 91	Resp: 370150
Ion Ratio	Lower Upper
91	100
92	51.3 25.8 77.3
134	24.7 13.5 40.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#90

Hexachloroethane

Concen: 53.086 ug/l




RT: 13.959 min Scan# 1991


Delta R.T. 0.001 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33


Tgt Ion:117	Resp: 78345
Ion Ratio	Lower Upper
117	100
201	81.2 41.6 124.8





#91
1,2-Dichlorobenzene
Concen: 54.063 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33




Instrument : MSVOA_Y
ClientSampleId : CB-14-G1-SO-14-050923MSD




Tgt Ion:146 Resp: 185174
Ion Ratio Lower Upper
146 100
111 40.2 19.8 59.4
148 63.8 31.8 95.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 05/16/2023
Supervised By :Mahesh Dadoda 05/17/2023




#92
1,2-Dibromo-3-Chloropropane
Concen: 54.546 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.000 min
Lab File: VY013747.D
Acq: 15 May 2023 19:33



Tgt Ion: 75 Resp: 19347
Ion Ratio Lower Upper
75 100
155 81.4 44.6 134.0
157 105.3 57.6 172.7





#93

1,2,4-Trichlorobenzene

Concen: 53.208 ug/l

RT: 15.007 min Scan# 2163

Delta R.T. 0.000 min

Lab File: VY013747.D


Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD



Tgt Ion:180 Resp: 108650




Ion	Ratio	Lower	Upper
180	100		
182	96.0	48.5	145.5
145	29.0	15.0	45.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#94

Hexachlorobutadiene

Concen: 55.492 ug/l

RT: 15.111 min Scan# 2180




Delta R.T. -0.000 min


Lab File: VY013747.D

Acq: 15 May 2023 19:33

Tgt Ion:225 Resp: 65950

Ion	Ratio	Lower	Upper
225	100		
223	62.9	31.4	94.2
227	64.0	32.0	96.0





#95

Naphthalene

Concen: 55.860 ug/l

RT: 15.233 min Scan# 2200

Delta R.T. -0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Instrument :

MSVOA_Y

ClientSampleId :

CB-14-G1-SO-14-050923MSD





Tgt Ion:128 Resp: 247500
 Ion Ratio Lower Upper
 128 100
 127 13.5 10.7 16.1
 129 11.1 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 05/16/2023

Supervised By :Mahesh Dadoda 05/17/2023



#96

1,2,3-Trichlorobenzene

Concen: 54.998 ug/l

RT: 15.422 min Scan# 2231

Delta R.T. -0.000 min

Lab File: VY013747.D

Acq: 15 May 2023 19:33

Tgt Ion:180 Resp: 100101
 Ion Ratio Lower Upper
 180 100
 182 95.6 47.1 141.3
 145 31.0 15.9 47.7

