

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
 Data File : VY013779.D
 Acq On : 17 May 2023 16:26
 Operator : KP/MD
 Sample : 02802-01RE
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ORA-1639RE

Quant Time: May 18 07:15:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 18 06:54:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	1421	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	2459	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	2026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	393	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	1523	92.262	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 184.520%#	
35) Dibromofluoromethane	7.722	113	823	52.193	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.380%	
50) Toluene-d8	10.179	98	3052	51.053	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.100%	
62) 4-Bromofluorobenzene	12.489	95	636	32.030	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 64.060%	
Target Compounds						Qvalue
44) Trichloroethene	8.941	130	236	14.259	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
Data File : VY013779.D
Acq On : 17 May 2023 16:26
Operator : KP/MD
Sample : 02802-01RE
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ORA-1639RE

Quant Time: May 18 07:15:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
Quant Title : SW846 8260
QLast Update : Thu May 18 06:54:18 2023
Response via : Initial Calibration

