

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
 Data File : VY013780.D
 Acq On : 17 May 2023 16:50
 Operator : KP/MD
 Sample : 02774-01
 Misc : 7.79g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-N-1-VOC

Quant Time: May 18 07:16:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 18 06:54:18 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	158755	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	280615	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	270053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	115517	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	110804	60.082	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.160%	
35) Dibromofluoromethane	7.722	113	97657	54.270	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.540%	
50) Toluene-d8	10.179	98	345012	50.573	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.140%	
62) 4-Bromofluorobenzene	12.483	95	109997	48.543	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 97.080%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.722	84	25330	10.218	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
Data File : VY013780.D
Acq On : 17 May 2023 16:50
Operator : KP/MD
Sample : 02774-01
Misc : 7.79g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-N-1-VOC

Quant Time: May 18 07:16:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
Quant Title : SW846 8260
QLast Update : Thu May 18 06:54:18 2023
Response via : Initial Calibration

