

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
 Data File : VY013788.D
 Acq On : 17 May 2023 19:57
 Operator : KP/MD
 Sample : 02812-15
 Misc : 4.79g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WG-0516-B2-0-2

Quant Time: May 18 07:18:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 18 06:54:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	164638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	281043	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	260081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	106523	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	114580	59.910	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.820%
35) Dibromofluoromethane	7.722	113	96297	53.433	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.860%
50) Toluene-d8	10.185	98	342907	50.188	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.380%
62) 4-Bromofluorobenzene	12.483	95	103715	45.701	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
Data File : VY013788.D
Acq On : 17 May 2023 19:57
Operator : KP/MD
Sample : 02812-15
Misc : 4.79g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WG-0516-B2-0-2

Quant Time: May 18 07:18:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
Quant Title : SW846 8260
QLast Update : Thu May 18 06:54:18 2023
Response via : Initial Calibration

