

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
 Data File : VY013774.D
 Acq On : 17 May 2023 13:45
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY051723

Quant Time: May 18 06:57:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 18 06:54:18 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.359	0.345	3.9	99	0.00
3 P	Chloromethane	0.544	0.531	2.4	103	0.00
4 C	Vinyl Chloride	0.543	0.538	0.9#	102	0.00
5 T	Bromomethane	0.482	0.432	10.4	103	0.00
6 T	Chloroethane	0.374	0.370	1.1	106	0.00
7 T	Trichlorofluoromethane	0.791	0.761	3.8	99	0.00
8 T	Diethyl Ether	0.265	0.257	3.0	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.463	0.448	3.2	101	-0.01
10 T	Methyl Iodide	0.509	0.523	-2.8	99	-0.01
11 T	Tert butyl alcohol	0.047	0.056	-19.1	120	-0.02
12 CM	1,1-Dichloroethene	0.420	0.410	2.4#	99	0.00
13 T	Acrolein	0.084	0.079	6.0	94	-0.01
14 T	Allyl chloride	0.827	0.834	-0.8	100	-0.01
15 T	Acrylonitrile	0.155	0.150	3.2	96	-0.01
16 T	Acetone	0.175	0.179	-2.3	100	0.00
17 T	Carbon Disulfide	1.206	1.147	4.9	100	-0.01
18 T	Methyl Acetate	0.611	0.574	6.1	95	-0.02
19 T	Methyl tert-butyl Ether	1.273	1.292	-1.5	98	-0.01
20 T	Methylene Chloride	0.662	0.564	14.8	104	-0.02
21 T	trans-1,2-Dichloroethene	0.472	0.468	0.8	100	-0.01
22 T	Diisopropyl ether	1.758	1.843	-4.8	101	0.00
23 T	Vinyl Acetate	1.017	1.077	-5.9	100	-0.01
24 P	1,1-Dichloroethane	0.961	0.960	0.1	101	-0.01
25 T	2-Butanone	0.228	0.233	-2.2	99	0.00
26 T	2,2-Dichloropropane	0.821	0.843	-2.7	103	0.00
27 T	cis-1,2-Dichloroethene	0.557	0.565	-1.4	100	-0.01
28 T	Bromochloromethane	0.336	0.314	6.5	99	-0.01
29 T	Tetrahydrofuran	0.136	0.136	0.0	96	0.00
30 C	Chloroform	0.969	0.964	0.5#	100	-0.01
31 T	Cyclohexane	0.833	0.780	6.4	101	0.00
32 T	1,1,1-Trichloroethane	0.827	0.825	0.2	101	-0.01
33 S	1,2-Dichloroethane-d4	0.581	0.566	2.6	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.321	0.321	0.0	100	0.00
36 T	1,1-Dichloropropene	0.443	0.439	0.9	100	0.00
37 T	Ethyl Acetate	0.305	0.305	0.0	99	-0.01
38 T	Carbon Tetrachloride	0.467	0.459	1.7	100	0.00
39 T	Methylcyclohexane	0.504	0.504	0.0	101	0.00
40 TM	Benzene	1.287	1.277	0.8	100	0.00
41 T	Methacrylonitrile	0.151	0.160	-6.0	119	0.00
42 TM	1,2-Dichloroethane	0.416	0.410	1.4	99	0.00
43 T	Isopropyl Acetate	0.553	0.558	-0.9	98	0.00
44 TM	Trichloroethene	0.337	0.331	1.8	100	0.00
45 C	1,2-Dichloropropane	0.354	0.359	-1.4#	101	0.00
46 T	Dibromomethane	0.198	0.193	2.5	97	0.00
47 T	Bromodichloromethane	0.469	0.471	-0.4	99	0.00
48 T	Methyl methacrylate	0.254	0.262	-3.1	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
 Data File : VY013774.D
 Acq On : 17 May 2023 13:45
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY051723

Quant Time: May 18 06:57:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 18 06:54:18 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	96	0.00
50 S	Toluene-d8	1.216	1.229	-1.1	101	0.00
51 T	4-Methyl-2-Pentanone	0.308	0.309	-0.3	95	0.00
52 CM	Toluene	0.780	0.792	-1.5#	100	0.00
53 T	t-1,3-Dichloropropene	0.484	0.492	-1.7	100	0.00
54 T	cis-1,3-Dichloropropene	0.540	0.550	-1.9	102	0.00
55 T	1,1,2-Trichloroethane	0.269	0.269	0.0	99	0.00
56 T	Ethyl methacrylate	0.360	0.380	-5.6	97	0.00
57 T	1,3-Dichloropropane	0.464	0.464	0.0	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.130	0.117	10.0	97	0.00
59 T	2-Hexanone	0.219	0.228	-4.1	97	0.00
60 T	Dibromochloromethane	0.327	0.323	1.2	97	0.00
61 T	1,2-Dibromoethane	0.254	0.256	-0.8	99	0.00
62 S	4-Bromofluorobenzene	0.404	0.417	-3.2	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.317	0.303	4.4	98	0.00
65 PM	Chlorobenzene	0.946	0.941	0.5	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.359	2.4	97	0.00
67 C	Ethyl Benzene	1.682	1.723	-2.4#	101	0.00
68 T	m/p-Xylenes	0.630	0.646	-2.5	100	0.00
69 T	o-Xylene	0.599	0.620	-3.5	101	0.00
70 T	Styrene	1.018	1.057	-3.8	101	0.00
71 P	Bromoform	0.243	0.246	-1.2	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.385	3.471	-2.5	101	0.00
74 T	N-amyl acetate	1.149	1.183	-3.0	98	0.00
75 P	1,1,2,2-Tetrachloroethane	0.802	0.799	0.4	98	0.00
76 T	1,2,3-Trichloropropane	0.605	0.561	7.3	93	0.00
77 T	Bromobenzene	0.817	0.816	0.1	100	0.00
78 T	n-propylbenzene	4.170	4.309	-3.3	101	0.00
79 T	2-Chlorotoluene	2.353	2.414	-2.6	102	0.00
80 T	1,3,5-Trimethylbenzene	2.760	2.846	-3.1	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.272	0.277	-1.8	98	0.00
82 T	4-Chlorotoluene	2.451	2.500	-2.0	102	0.00
83 T	tert-Butylbenzene	2.404	2.521	-4.9	103	0.00
84 T	1,2,4-Trimethylbenzene	2.705	2.815	-4.1	102	0.00
85 T	sec-Butylbenzene	3.627	3.777	-4.1	103	0.00
86 T	p-Isopropyltoluene	2.928	3.061	-4.5	103	0.00
87 T	1,3-Dichlorobenzene	1.590	1.603	-0.8	102	0.00
88 T	1,4-Dichlorobenzene	1.598	1.594	0.3	101	0.00
89 T	n-Butylbenzene	2.871	3.014	-5.0	104	0.00
90 T	Hexachloroethane	0.619	0.608	1.8	101	0.00
91 T	1,2-Dichlorobenzene	1.416	1.435	-1.3	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.130	0.124	4.6	94	0.00
93 T	1,2,4-Trichlorobenzene	0.840	0.861	-2.5	105	0.00
94 T	Hexachlorobutadiene	0.522	0.515	1.3	105	0.00
95 T	Naphthalene	1.667	1.772	-6.3	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
Data File : VY013774.D
Acq On : 17 May 2023 13:45
Operator : KP/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY051723

Quant Time: May 18 06:57:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
Quant Title : SW846 8260
QLast Update : Thu May 18 06:54:18 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.756	0.770	-1.9	106	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6