



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051723\
Data File : VY013766.D
Acq On : 17 May 2023 09:11
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Title : SW846 8260
Last Update : Tue May 16 04:07:42 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.9	23109	PASS
75	95	30	60	53.2	53691	PASS
95	95	100	100	100.0	100896	PASS
96	95	5	9	7.0	7059	PASS
173	174	0.00	2	1.8	1584	PASS
174	95	50	100	85.2	85981	PASS
175	174	5	9	7.8	6740	PASS
176	174	95	101	95.5	82107	PASS
177	176	5	9	6.7	5462	PASS

Average of 12.477 to 12.489 min.: VY013766.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
36.00	1205	46.00	181	55.95	1935	66.95	374
37.00	6236	46.20	57	57.00	3469	68.00	1129
38.00	5865	47.05	1727	58.05	122	69.00	1096
39.05	2633	47.95	764	58.95	122	70.00	99
39.95	195	49.00	4982	59.95	1146	70.90	76
41.00	136	50.00	23109	61.00	5848	71.95	641
41.90	68	51.05	6957	62.05	5798	73.00	5635
42.80	26	52.00	324	63.00	4852	74.00	19283
43.10	124	52.90	179	64.05	810	75.00	53691
44.00	936	53.90	19	65.00	878	76.00	4566
45.00	1132	54.90	403	65.90	117	76.95	641

Average of 12.477 to 12.489 min.: VY013766.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
77.80	492	88.70	86	105.90	514	115.50	58
78.90	4728	90.80	584	106.75	164	115.95	415
79.85	1496	92.00	3424	108.00	59	116.90	893
80.90	5194	92.95	4967	108.90	20	117.90	515
81.90	1425	94.00	12891	109.80	121	118.90	872
82.70	50	95.00	100896	110.10	21	123.65	82
83.00	136	96.00	7059	110.95	176	124.10	22
85.80	184	97.05	181	111.70	43	125.00	166
86.00	80	102.95	197	111.95	65	125.20	74
86.95	4705	103.95	612	112.90	111	125.60	76
87.95	4407	105.05	184	114.95	182	126.10	52

Average of 12.477 to 12.489 min.: VY013766.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
127.90	320	139.95	73	147.95	311	164.85	44
128.95	216	140.90	1418	148.80	111	170.55	125
129.90	419	141.70	59	149.80	142	170.90	72
130.80	33	141.95	154	152.85	83	171.20	80
131.00	154	142.90	1798	153.95	77	171.95	1177
132.75	230	143.70	26	154.85	325	173.00	1584
133.15	402	143.95	74	156.95	193	173.90	85981
133.85	111	144.95	542	158.80	129	174.90	6740
134.70	80	145.60	56	159.10	64	175.90	82107
134.95	208	145.95	205	160.70	196	176.90	5462
136.90	211	146.90	184	162.95	56	177.85	200

Average of 12.477 to 12.489 min.: VY013766.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
178.95	193	207.90	23	267.00	129		
189.10	69	235.00	21	279.20	28		
190.80	95	248.70	24	281.00	4075		
191.05	342	248.85	166	282.10	1255		
191.80	68	249.20	31	282.95	985		
192.90	759	250.00	69	284.10	152		
194.10	121	250.95	153	284.80	59		
194.90	67	252.90	42				
203.10	36	264.70	78				
204.85	121	264.90	322				
206.95	450	266.05	164				