

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051823\
 Data File : VY013807.D
 Acq On : 18 May 2023 18:03
 Operator : KP/MD
 Sample : 02832-01
 Misc : 5.75g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-22-K2-SO-64-051723

Quant Time: May 19 03:23:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 18 06:54:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	170620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	280338	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	265945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	119211	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	115108	58.076	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.160%
35) Dibromofluoromethane	7.722	113	92599	51.510	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.020%
50) Toluene-d8	10.185	98	342984	50.326	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.660%
62) 4-Bromofluorobenzene	12.483	95	106103	46.871	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051823\
Data File : VY013807.D
Acq On : 18 May 2023 18:03
Operator : KP/MD
Sample : 02832-01
Misc : 5.75g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CB-22-K2-SO-64-051723

Quant Time: May 19 03:23:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051723S.M
Quant Title : SW846 8260
QLast Update : Thu May 18 06:54:18 2023
Response via : Initial Calibration

