

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052022\
 Data File : VY008923.D
 Acq On : 20 May 2022 23:41
 Operator : KP/MD
 Sample : N2908-01
 Misc : 5.71g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BI-1

Quant Time: May 21 02:48:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	142736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	250294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	237410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	103319	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	88966	60.562	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.120%
35) Dibromofluoromethane	7.716	113	79765	51.249	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.500%
50) Toluene-d8	10.179	98	312359	54.400	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.800%
62) 4-Bromofluorobenzene	12.483	95	94141	40.644	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052022\
Data File : VY008923.D
Acq On : 20 May 2022 23:41
Operator : KP/MD
Sample : N2908-01
Misc : 5.71g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BI-1

Quant Time: May 21 02:48:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

