

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052022\
 Data File : VY008912.D
 Acq On : 20 May 2022 18:39
 Operator : KP/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/24/2022
 Supervised By : Mahesh Dadoda 05/24/2022

Quant Time: May 20 22:59:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 22:58:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	127204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	208319	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	192654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	92753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	7575	6.665	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	13.320%#		
35) Dibromofluoromethane	7.716	113	6728	5.665	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	11.340%#		
50) Toluene-d8	10.179	98	25669	5.768	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	11.540%#		
62) 4-Bromofluorobenzene	12.483	95	8369	4.945	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	9.900%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.900	85	6140	5.962	ug/l	99
3) Chloromethane	2.113	50	9941	6.339	ug/l #	88
4) Vinyl Chloride	2.247	62	13764	6.183	ug/l	94
5) Bromomethane	2.643	94	13045	7.550	ug/l	100
6) Chloroethane	2.796	64	9567	5.914	ug/l	93
7) Trichlorofluoromethane	3.125	101	10973	5.358	ug/l	100
8) Diethyl Ether	3.521	74	3472	5.925	ug/l	85
9) 1,1,2-Trichlorotrifluo...	3.899	101	7072	5.749	ug/l	90
10) Methyl Iodide	4.094	142	4773m	3.482	ug/l	
11) Tert butyl alcohol	4.948	59	3810	28.932	ug/l #	95
12) 1,1-Dichloroethene	3.869	96	6158	5.490	ug/l #	74
13) Acrolein	3.735	56	1565	25.441	ug/l	98
14) Allyl chloride	4.485	41	9083	5.466	ug/l #	90
15) Acrylonitrile	5.161	53	9414	24.566	ug/l	97
16) Acetone	3.948	43	8474	27.913	ug/l	96
17) Carbon Disulfide	4.192	76	19671	5.773	ug/l	97
18) Methyl Acetate	4.478	43	4747	4.898	ug/l	95
19) Methyl tert-butyl Ether	5.222	73	17579	5.741	ug/l	99
20) Methylene Chloride	4.716	84	16510	10.320	ug/l #	82
21) trans-1,2-Dichloroethene	5.216	96	7592	5.507	ug/l	85
22) Diisopropyl ether	6.118	45	20452	5.727	ug/l	97
23) Vinyl Acetate	6.057	43	66053	26.543	ug/l #	88
24) 1,1-Dichloroethane	6.015	63	13099	5.628	ug/l	98
25) 2-Butanone	6.984	43	12991	25.985	ug/l #	86
26) 2,2-Dichloropropane	6.978	77	12559	5.722	ug/l	100
27) cis-1,2-Dichloroethene	6.978	96	8314	5.402	ug/l	90
28) Bromochloromethane	7.338	49	4604	5.401	ug/l	90
29) Tetrahydrofuran	7.350	42	8431	25.225	ug/l #	86
30) Chloroform	7.502	83	14309	5.746	ug/l	96
31) Cyclohexane	7.783	56	14293	6.491	ug/l #	76
32) 1,1,1-Trichloroethane	7.697	97	12402	5.372	ug/l	98
36) 1,1-Dichloropropene	7.911	75	11384	5.620	ug/l	99
37) Ethyl Acetate	7.069	43	6330	5.377	ug/l	97
38) Carbon Tetrachloride	7.899	117	10995	5.087	ug/l	94
39) Methylcyclohexane	9.179	83	12906	5.264	ug/l #	85
40) Benzene	8.161	78	30817	5.420	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052022\
 Data File : VY008912.D
 Acq On : 20 May 2022 18:39
 Operator : KP/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/24/2022
 Supervised By : Mahesh Dadoda 05/24/2022

Quant Time: May 20 22:59:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 22:58:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.313	41	2704m	4.361	ug/l	
42) 1,2-Dichloroethane	8.240	62	9590	5.744	ug/l	94
43) Isopropyl Acetate	8.270	43	10417	5.224	ug/l #	89
44) Trichloroethene	8.941	130	8331	5.417	ug/l	100
45) 1,2-Dichloropropane	9.215	63	7516	5.543	ug/l	97
46) Dibromomethane	9.301	93	4780	5.385	ug/l	87
47) Bromodichloromethane	9.490	83	10951	5.559	ug/l	87
48) Methyl methacrylate	9.295	41	4230	4.792	ug/l #	74
49) 1,4-Dioxane	9.301	88	1302	100.785	ug/l #	80
51) 4-Methyl-2-Pentanone	10.069	43	28039	24.064	ug/l	91
52) Toluene	10.246	92	19671	5.056	ug/l	96
53) t-1,3-Dichloropropene	10.465	75	10455	5.079	ug/l	91
54) cis-1,3-Dichloropropene	9.929	75	11997	5.413	ug/l #	88
55) 1,1,2-Trichloroethane	10.642	97	6282	5.205	ug/l	97
56) Ethyl methacrylate	10.508	69	7097	4.780	ug/l	91
57) 1,3-Dichloropropane	10.788	76	10686	5.375	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.782	63	17052	23.416	ug/l	93
59) 2-Hexanone	10.831	43	17837	21.625	ug/l	92
60) Dibromochloromethane	10.983	129	7350	5.040	ug/l	100
61) 1,2-Dibromoethane	11.087	107	6188	5.132	ug/l	96
64) Tetrachloroethene	10.715	164	6945	5.418	ug/l	93
65) Chlorobenzene	11.514	112	21265	5.269	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.587	131	7395	4.993	ug/l	95
67) Ethyl Benzene	11.593	91	36143	4.983	ug/l	99
68) m/p-Xylenes	11.703	106	26836	9.231	ug/l	97
69) o-Xylene	12.032	106	13120	4.889	ug/l	92
70) Styrene	12.044	104	20659	4.512	ug/l	93
71) Bromoform	12.209	173	4483	4.461	ug/l #	95
73) Isopropylbenzene	12.331	105	33206	5.062	ug/l	100
74) N-amyl acetate	12.142	43	8492	5.456	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.581	83	8174	5.681	ug/l	98
76) 1,2,3-Trichloropropane	12.629	75	4613m	4.247	ug/l	
77) Bromobenzene	12.611	156	8294	5.459	ug/l	95
78) n-propylbenzene	12.672	91	41899	5.572	ug/l	99
79) 2-Chlorotoluene	12.757	91	22546	5.161	ug/l	92
80) 1,3,5-Trimethylbenzene	12.812	105	26894	5.048	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.380	75	2467	5.001	ug/l	97
82) 4-Chlorotoluene	12.855	91	24030	5.189	ug/l	99
83) tert-Butylbenzene	13.074	119	24185	5.067	ug/l	93
84) 1,2,4-Trimethylbenzene	13.123	105	26491	5.014	ug/l	99
85) sec-Butylbenzene	13.257	105	37548	5.048	ug/l	99
86) p-Isopropyltoluene	13.373	119	28839	4.745	ug/l	99
87) 1,3-Dichlorobenzene	13.367	146	16685	5.212	ug/l	98
88) 1,4-Dichlorobenzene	13.446	146	17729	5.603	ug/l	97
89) n-Butylbenzene	13.696	91	28583	4.992	ug/l	99
90) Hexachloroethane	13.965	117	6192	5.399	ug/l	84
91) 1,2-Dichlorobenzene	13.739	146	14995	5.372	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.355	75	1219	5.024	ug/l	87
93) 1,2,4-Trichlorobenzene	15.007	180	8291	5.444	ug/l	97
94) Hexachlorobutadiene	15.117	225	4710	5.411	ug/l	95
95) Naphthalene	15.233	128	15735	4.697	ug/l	98
96) 1,2,3-Trichlorobenzene	15.428	180	7639	5.726	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052022\
Data File : VY008912.D
Acq On : 20 May 2022 18:39
Operator : KP/MD
Sample : VSTDICC005
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022

Quant Time: May 20 22:59:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 22:58:02 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052022\
Data File : VY008912.D
Acq On : 20 May 2022 18:39
Operator : KP/MD
Sample : VSTDICC005
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1


Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/24/2022
Supervised By : Mahesh Dadoda 05/24/2022


Quant Time: May 20 22:59:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 22:58:02 2022
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion:168 Resp: 127204
Ion Ratio Lower Upper
168 100
99 52.6 46.9 70.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#2
Dichlorodifluoromethane
Concen: 5.962 ug/l
RT: 1.900 min Scan# 13
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Tgt Ion: 85 Resp: 6140
Ion Ratio Lower Upper
85 100
87 31.5 16.1 48.2





#3
Chloromethane
Concen: 6.339 ug/l
RT: 2.113 min Scan# 41
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion: 50 Resp: 994
Ion Ratio Lower Upper
50 100
52 37.3 24.6 37.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#4
Vinyl Chloride
Concen: 6.183 ug/l
RT: 2.247 min Scan# 70
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Tgt Ion: 62 Resp: 13764
Ion Ratio Lower Upper
62 100
64 34.9 25.0 37.6





#5
Bromomethane
Concen: 7.550 ug/l
RT: 2.643 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 94 Resp: 1304
Ion Ratio Lower Upper
94 100
96 91.6 73.5 110.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#6
Chloroethane
Concen: 5.914 ug/l
RT: 2.796 min Scan# 160
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Tgt Ion: 64 Resp: 9567
Ion Ratio Lower Upper
64 100
66 35.5 25.4 38.0





#7
Trichlorofluoromethane
Concen: 5.358 ug/l
RT: 3.125 min Scan# 213
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:101 Resp: 10972
Ion Ratio Lower Upper
101 100
103 60.0 48.0 72.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#8
Diethyl Ether
Concen: 5.925 ug/l
RT: 3.521 min Scan# 279
Delta R.T. -0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Tgt Ion: 74 Resp: 3472
Ion Ratio Lower Upper
74 100
45 94.7 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.749 ug/l
RT: 3.899 min Scan# 341
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:101 Resp: 7072
Ion Ratio Lower Upper
101 100
85 46.3 40.6 60.8
151 70.3 64.9 97.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#10
Methyl Iodide
Concen: 3.482 ug/l m
RT: 4.094 min Scan# 373
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion:142 Resp: 4773
Ion Ratio Lower Upper
142 100
127 46.7 41.5 62.3
141 6.9 11.9 17.9#





#11
Tert butyl alcohol
Concen: 28.932 ug/l
RT: 4.948 min Scan# 513
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 59 Resp: 3810
Ion Ratio Lower Upper
59 100
57 9.6 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#12
1,1-Dichloroethene
Concen: 5.490 ug/l
RT: 3.869 min Scan# 336
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Tgt Ion: 96 Resp: 6158
Ion Ratio Lower Upper
96 100
61 133.2 145.8 218.6#
98 67.3 52.5 78.7





#13
Acrolein
Concen: 25.441 ug/l
RT: 3.735 min Scan# 312
Delta R.T. 0.012 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 56 Resp: 156
Ion Ratio Lower Upper
56 100
55 70.3 57.6 86.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#14
Allyl chloride
Concen: 5.466 ug/l
RT: 4.485 min Scan# 437
Delta R.T. 0.012 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Tgt Ion: 41 Resp: 9083
Ion Ratio Lower Upper
41 100
39 78.5 58.8 88.2
76 42.3 25.1 37.7#





#15

Acrylonitrile

Concen: 24.566 ug/l

RT: 5.161 min Scan# 54

Delta R.T. 0.006 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005





Tgt Ion: 53	Resp: 9414
Ion Ratio	Lower Upper
53	100
52	81.3 67.0 100.4
51	39.2 29.8 44.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#16

Acetone

Concen: 27.913 ug/l




RT: 3.948 min Scan# 349


Delta R.T. -0.000 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39

Tgt Ion: 43	Resp: 8474
Ion Ratio	Lower Upper
43	100
58	29.7 22.0 33.0









#17
Carbon Disulfide
Concen: 5.773 ug/l
RT: 4.192 min Scan# 31
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 76 Resp: 1967
Ion Ratio Lower Upper
76 100
78 10.3 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#18
Methyl Acetate
Concen: 4.898 ug/l
RT: 4.478 min Scan# 436
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion: 43 Resp: 4747
Ion Ratio Lower Upper
43 100
74 22.9 16.6 24.8





#19

Methyl tert-butyl Ether

Concen: 5.741 ug/l

RT: 5.222 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 73 Resp: 17579

Ion Ratio Lower Upper

73 100





57 22.6 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#20

Methylene Chloride

Concen: 10.320 ug/l

RT: 4.716 min Scan# 475

Delta R.T. 0.006 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39

Tgt Ion: 84 Resp: 16510




Ion Ratio Lower Upper


84 100

49 113.0 114.4 171.6#

51 35.6 35.3 52.9


86 60.1 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 5.507 ug/l
RT: 5.216 min Scan# 51
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 96 Resp: 7592
Ion Ratio Lower Upper
96 100
61 125.1 120.5 180.7
98 63.9 49.8 74.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#22
Diisopropyl ether
Concen: 5.727 ug/l
RT: 6.118 min Scan# 705
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Tgt Ion: 45 Resp: 20452
Ion Ratio Lower Upper
45 100
43 64.9 49.3 73.9
87 24.8 19.2 28.8
59 11.6 8.4 12.6






#23
 Vinyl Acetate
 Concen: 26.543 ug/l
 RT: 6.057 min Scan# 695
 Delta R.T. -0.000 min
 Lab File: VY008912.D
 Acq: 20 May 2022 18:39

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005






Tgt Ion: 43 Resp: 6605
 Ion Ratio Lower Upper
 43 100
 86 13.3 7.1 10.7


Manual Integrations

APPROVED



Reviewed By :John Carlone 05/24/2022
 Supervised By :Mahesh Dadoda 05/24/2022




#24
 1,1-Dichloroethane
 Concen: 5.628 ug/l
 RT: 6.015 min Scan# 688
 Delta R.T. -0.000 min
 Lab File: VY008912.D
 Acq: 20 May 2022 18:39



Tgt Ion: 63 Resp: 13099
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.6 10.8
 100 2.7 2.0 5.8









#25
2-Butanone
Concen: 25.985 ug/l
RT: 6.984 min Scan# 847
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument : MSVOA_Y
Client Sample Id : VSTDIC005

Tgt Ion: 43 Resp: 1299
Ion Ratio Lower Upper
43 100
72 28.7 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#26
2,2-Dichloropropane
Concen: 5.722 ug/l
RT: 6.978 min Scan# 846
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion: 77 Resp: 12559
Ion Ratio Lower Upper
77 100
97 21.0 10.4 31.4





#27
 cis-1,2-Dichloroethene
 Concen: 5.402 ug/l
 RT: 6.978 min Scan# 84
 Delta R.T. -0.006 min
 Lab File: VY008912.D
 Acq: 20 May 2022 18:39




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005




Tgt Ion: 96 Resp: 8314
 Ion Ratio Lower Upper
 96 100
 61 145.9 0.0 319.8
 98 70.2 0.0 127.0

Manual Integrations
 APPROVED



Reviewed By :John Carlone 05/24/2022
 Supervised By :Mahesh Dadoda 05/24/2022




#28
 Bromochloromethane
 Concen: 5.401 ug/l
 RT: 7.338 min Scan# 905
 Delta R.T. 0.006 min
 Lab File: VY008912.D
 Acq: 20 May 2022 18:39



Tgt Ion: 49 Resp: 4604
 Ion Ratio Lower Upper
 49 100
 129 0.5 0.0 2.2
 130 77.8 55.4 83.0









#29
Tetrahydrofuran
Concen: 25.225 ug/l
RT: 7.350 min Scan# 906
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 42 Resp: 843
Ion Ratio Lower Upper
42 100
72 45.9 30.3 45.5
71 44.5 29.0 43.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#30
Chloroform
Concen: 5.746 ug/l
RT: 7.502 min Scan# 932
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion: 83 Resp: 14309
Ion Ratio Lower Upper
83 100
85 61.9 51.8 77.6





#31
Cyclohexane
Concen: 6.491 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 56 Resp: 1429
Ion Ratio Lower Upper
56 100
69 52.8 26.7 40.1
84 101.9 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#32
1,1,1-Trichloroethane
Concen: 5.372 ug/l
RT: 7.697 min Scan# 964
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Tgt Ion: 97 Resp: 12402
Ion Ratio Lower Upper
97 100
99 62.1 50.7 76.1
61 48.5 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 6.665 ug/l

RT: 8.142 min Scan# 1037

Delta R.T. -0.000 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 757

Ion Ratio Lower Upper

65 100





67 54.9 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39




Tgt Ion:114 Resp: 208319


Ion Ratio Lower Upper

114 100

63 18.4 0.0 43.2


88 13.8 0.0 31.8





#35
Dibromofluoromethane
Concen: 5.665 ug/l
RT: 7.716 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:113 Resp: 6728
Ion Ratio Lower Upper
113 100
111 103.8 82.6 124.0
192 19.4 17.6 26.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#36
1,1-Dichloropropene
Concen: 5.620 ug/l
RT: 7.911 min Scan# 999
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Tgt Ion: 75 Resp: 11384
Ion Ratio Lower Upper
75 100
110 33.0 17.1 51.2
77 30.8 24.4 36.6





#37
Ethyl Acetate
Concen: 5.377 ug/l
RT: 7.069 min Scan# 861
Delta R.T. -0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 43 Resp: 6330
Ion Ratio Lower Upper
43 100
61 13.4 9.7 14.5
70 9.4 7.1 10.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#38
Carbon Tetrachloride
Concen: 5.087 ug/l
RT: 7.899 min Scan# 997
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Tgt Ion:117 Resp: 10995
Ion Ratio Lower Upper
117 100
119 104.8 77.9 116.9
121 30.2 24.8 37.2





#39

Methylcyclohexane

Concen: 5.264 ug/l

RT: 9.179 min Scan# 1207

Delta R.T. -0.000 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 83 Resp: 12900





Ion	Ratio	Lower	Upper
83	100		
55	64.2	66.5	99.7
98	43.8	37.9	56.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#40

Benzene

Concen: 5.420 ug/l

RT: 8.161 min Scan# 1040




Delta R.T. -0.000 min


Lab File: VY008912.D

Acq: 20 May 2022 18:39

Tgt Ion: 78 Resp: 30817


Ion	Ratio	Lower	Upper
78	100		
77	24.8	20.4	30.6





#41
Methacrylonitrile
Concen: 4.361 ug/l m
RT: 7.313 min Scan# 901
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion: 41 Resp: 2704
Ion Ratio Lower Upper
41 100
39 181.7 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#42
1,2-Dichloroethane
Concen: 5.744 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion: 62 Resp: 9590
Ion Ratio Lower Upper
62 100
98 9.6 0.0 14.8





#43
Isopropyl Acetate
Concen: 5.224 ug/l
RT: 8.270 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument : MSVOA_Y
ClientSampleId : VSTDIC005



Tgt Ion: 43 Resp: 1041
Ion Ratio Lower Upper
43 100
61 31.0 20.2 30.2
87 14.2 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#44
Trichloroethene
Concen: 5.417 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion:130 Resp: 8331
Ion Ratio Lower Upper
130 100
95 94.0 0.0 188.2





#45

1,2-Dichloropropane

Concen: 5.543 ug/l

RT: 9.215 min Scan# 1213

Delta R.T. -0.000 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 7510

Ion Ratio Lower Upper

63 100





65 27.1 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#46

Dibromomethane

Concen: 5.385 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. -0.000 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39




Tgt Ion: 93 Resp: 4780


Ion Ratio Lower Upper

93 100

95 90.0 64.3 96.5

174 89.8 83.4 125.0





#47

Bromodichloromethane

Concen: 5.559 ug/l

RT: 9.490 min Scan# 1259

Delta R.T. -0.006 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005





Tgt Ion:	83	Resp:	1095
Ion	Ratio	Lower	Upper
83	100		
85	53.7	51.9	77.9
127	7.5	6.6	10.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#48

Methyl methacrylate

Concen: 4.792 ug/l




RT: 9.295 min Scan# 1226


Delta R.T. 0.006 min

Lab File: VY008912.D

Acq: 20 May 2022 18:39


Tgt Ion:	41	Resp:	4230
Ion	Ratio	Lower	Upper
41	100		
69	97.7	58.2	87.2#
39	75.5	47.4	71.0#





#49
1,4-Dioxane
Concen: 100.785 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. 0.012 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion: 88 Resp: 1302
Ion Ratio Lower Upper
88 100
43 37.2 32.8 49.2
58 51.3 59.8 89.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#50
Toluene-d8
Concen: 5.768 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion: 98 Resp: 25669
Ion Ratio Lower Upper
98 100
100 64.3 50.2 75.2





#51
4-Methyl-2-Pentanone
Concen: 24.064 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 43 Resp: 28039
Ion Ratio Lower Upper
43 100
58 40.2 28.1 42.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#52
Toluene
Concen: 5.056 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Tgt Ion: 92 Resp: 19671
Ion Ratio Lower Upper
92 100
91 172.6 142.6 213.8









#53
t-1,3-Dichloropropene
Concen: 5.079 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument : MSVOA_Y
ClientSampleId : VSTDIC005

Tgt Ion: 75 Resp: 1045
Ion Ratio Lower Upper
75 100
77 37.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#54
cis-1,3-Dichloropropene
Concen: 5.413 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion: 75 Resp: 11997
Ion Ratio Lower Upper
75 100
77 34.6 25.6 38.4
39 47.2 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 5.205 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion	Ratio	Lower	Upper
97	100		
83	86.5	69.8	104.8
85	56.6	44.3	66.5
99	56.0	49.8	74.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#56
Ethyl methacrylate
Concen: 4.780 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Tgt Ion	Ratio	Lower	Upper
69	100		
41	68.2	63.4	95.0
39	42.9	35.8	53.8





#57
1,3-Dichloropropane
Concen: 5.375 ug/l
RT: 10.788 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument : MSVOA_Y
ClientSampleId : VSTDIC005



Tgt Ion: 76 Resp: 10680
Ion Ratio Lower Upper
76 100
78 33.1 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#58
2-Chloroethyl Vinyl ether
Concen: 23.416 ug/l
RT: 9.782 min Scan# 1306
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion: 63 Resp: 17052
Ion Ratio Lower Upper
63 100
106 30.4 21.5 32.3









#59
2-Hexanone
Concen: 21.625 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 1783
Ion Ratio Lower Upper
43 100
58 54.3 24.4 73.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#60
Dibromochloromethane
Concen: 5.040 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion:129 Resp: 7350
Ion Ratio Lower Upper
129 100
127 77.0 38.7 116.1









#61
1,2-Dibromoethane
Concen: 5.132 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:107 Resp: 6188
Ion Ratio Lower Upper
107 100
109 90.4 75.4 113.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#62
4-Bromofluorobenzene
Concen: 4.945 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion: 95 Resp: 8369
Ion Ratio Lower Upper
95 100
174 85.7 0.0 177.0
176 86.0 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion:117 Resp: 192654
Ion Ratio Lower Upper
117 100
82 54.7 43.0 64.4
119 32.7 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#64
Tetrachloroethene
Concen: 5.418 ug/l
RT: 10.715 min Scan# 1459
Delta R.T. -0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Tgt Ion:164 Resp: 6945
Ion Ratio Lower Upper
164 100
166 135.9 100.4 150.6
129 94.1 71.4 107.0
131 94.4 70.8 106.2





#65
Chlorobenzene
Concen: 5.269 ug/l
RT: 11.514 min Scan# 1159
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:112 Resp: 2126
Ion Ratio Lower Upper
112 100
114 30.3 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 4.993 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Tgt Ion:131 Resp: 7395
Ion Ratio Lower Upper
131 100
133 96.1 47.9 143.6
119 70.3 30.8 92.3





#67
Ethyl Benzene
Concen: 4.983 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 91 Resp: 3614
Ion Ratio Lower Upper
91 100
106 30.2 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#68
m/p-Xylenes
Concen: 9.231 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Tgt Ion:106 Resp: 26836
Ion Ratio Lower Upper
106 100
91 203.1 166.6 249.8





#69
o-Xylene
Concen: 4.889 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion:106 Resp: 13120
Ion Ratio Lower Upper
106 100
91 207.8 110.3 331.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#70
Styrene
Concen: 4.512 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion:104 Resp: 20659
Ion Ratio Lower Upper
104 100
78 45.8 42.4 63.6
103 53.2 44.4 66.6





#71
Bromoform
Concen: 4.461 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:173 Resp: 448
Ion Ratio Lower Upper
173 100
175 46.2 24.6 74.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Tgt Ion:152 Resp: 92753
Ion Ratio Lower Upper
152 100
115 57.3 28.2 84.7
150 160.7 0.0 347.6









#73
Isopropylbenzene
Concen: 5.062 ug/l
RT: 12.331 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:105 Resp: 33200
Ion Ratio Lower Upper
105 100
120 26.9 13.4 40.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#74
N-ethyl acetate
Concen: 5.456 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion: 43 Resp: 8492
Ion Ratio Lower Upper
43 100
70 42.9 28.7 43.1
55 27.9 17.3 25.9#
61 24.1 17.0 25.6









#75
1,1,2,2-Tetrachloroethane
Concen: 5.681 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 8174
Ion Ratio Lower Upper
83 100
131 9.8 5.6 16.8
85 66.2 32.3 96.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#76
1,2,3-Trichloropropane
Concen: 4.247 ug/l m
RT: 12.629 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion: 75 Resp: 4613
Ion Ratio Lower Upper
75 100
77 327.8 0.0 0.0#





#77
Bromobenzene
Concen: 5.459 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:156 Resp: 8294
Ion Ratio Lower Upper
156 100
77 182.3 95.5 286.4
158 93.6 48.0 144.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#78
n-propylbenzene
Concen: 5.572 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Tgt Ion: 91 Resp: 41899
Ion Ratio Lower Upper
91 100
120 23.2 11.3 33.9









#79
2-Chlorotoluene
Concen: 5.161 ug/l
RT: 12.757 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 22540
Ion Ratio Lower Upper
91 100
126 38.2 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#80
1,3,5-Trimethylbenzene
Concen: 5.048 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion:105 Resp: 26894
Ion Ratio Lower Upper
105 100
120 51.5 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 5.001 ug/l
RT: 12.380 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion: 75 Resp: 246
Ion Ratio Lower Upper
75 100
53 103.5 85.3 127.9
89 43.5 37.0 55.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#82
4-Chlorotoluene
Concen: 5.189 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39



Tgt Ion: 91 Resp: 24030
Ion Ratio Lower Upper
91 100
126 34.4 17.0 51.0









#83
tert-Butylbenzene
Concen: 5.067 ug/l
RT: 13.074 min Scan# 1846
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:119 Resp: 2418
Ion Ratio Lower Upper
119 100
91 61.9 34.8 104.4
134 27.5 13.3 39.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#84
1,2,4-Trimethylbenzene
Concen: 5.014 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39


Tgt Ion:105 Resp: 26491
Ion Ratio Lower Upper
105 100
120 45.2 22.3 66.9





#85
 sec-Butylbenzene
 Concen: 5.048 ug/l
 RT: 13.257 min Scan# 1875
 Delta R.T. 0.006 min
 Lab File: VY008912.D
 Acq: 20 May 2022 18:39




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005




Tgt Ion:105 Resp: 37548
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.2 30.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 05/24/2022
 Supervised By :Mahesh Dadoda 05/24/2022




#86
 p-Isopropyltoluene
 Concen: 4.745 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.006 min
 Lab File: VY008912.D
 Acq: 20 May 2022 18:39



Tgt Ion:119 Resp: 28839
 Ion Ratio Lower Upper
 119 100
 134 26.6 13.2 39.5
 91 25.2 12.4 37.2









#87
1,3-Dichlorobenzene
Concen: 5.212 ug/l
RT: 13.367 min Scan# 1191
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:146 Resp: 1668
Ion Ratio Lower Upper
146 100
111 39.1 19.5 58.5
148 61.4 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#88
1,4-Dichlorobenzene
Concen: 5.603 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion:146 Resp: 17729
Ion Ratio Lower Upper
146 100
111 41.9 19.3 57.8
148 63.6 32.6 97.8









#89
n-Butylbenzene
Concen: 4.992 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 2858
Ion Ratio Lower Upper
91 100
92 52.0 26.3 78.9
134 24.7 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#90
Hexachloroethane
Concen: 5.399 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion:117 Resp: 6192
Ion Ratio Lower Upper
117 100
201 71.5 43.0 128.8









#91
1,2-Dichlorobenzene
Concen: 5.372 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:146 Resp: 1499
Ion Ratio Lower Upper
146 100
111 39.0 20.3 60.9
148 61.3 32.3 96.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.024 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion: 75 Resp: 1219
Ion Ratio Lower Upper
75 100
155 85.5 38.0 113.9
157 109.7 47.8 143.3









#93
1,2,4-Trichlorobenzene
Concen: 5.444 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:180 Resp: 829
Ion Ratio Lower Upper
180 100
182 91.9 47.3 141.9
145 27.6 14.6 43.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#94
Hexachlorobutadiene
Concen: 5.411 ug/l
RT: 15.117 min Scan# 2181
Delta R.T. 0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion:225 Resp: 4710
Ion Ratio Lower Upper
225 100
223 69.1 30.8 92.3
227 62.9 31.6 94.7









#95
Naphthalene
Concen: 4.697 ug/l
RT: 15.233 min Scan# 2199
Delta R.T. -0.006 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:128 Resp: 15739
Ion Ratio Lower Upper
128 100
127 14.1 11.0 16.4
129 11.7 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#96
1,2,3-Trichlorobenzene
Concen: 5.726 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. -0.000 min
Lab File: VY008912.D
Acq: 20 May 2022 18:39

Tgt Ion:180 Resp: 7639
Ion Ratio Lower Upper
180 100
182 89.4 47.3 141.9
145 31.0 15.3 45.9

