

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052022\
 Data File : VY008929.D
 Acq On : 21 May 2022 02:02
 Operator : KP/MD
 Sample : N2916-03
 Misc : 6.40g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PP-1

Quant Time: May 21 02:49:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	133398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	230210	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	215060	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	93383	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	79258	57.730	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.460%
35) Dibromofluoromethane	7.716	113	71741	50.115	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.220%
50) Toluene-d8	10.179	98	286938	54.332	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.660%
62) 4-Bromofluorobenzene	12.483	95	90758	42.602	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.200%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052022\
Data File : VY008929.D
Acq On : 21 May 2022 02:02
Operator : KP/MD
Sample : N2916-03
Misc : 6.40g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PP-1

Quant Time: May 21 02:49:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008929.D
Acq: 21 May 2022 02:02




Instrument :
MSVOA_Y
ClientSampleId :
PP-1


Tgt Ion:168 Resp: 133398
Ion Ratio Lower Upper
168 100
99 50.3 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 57.730 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. -0.007 min
Lab File: VY008929.D
Acq: 21 May 2022 02:02

Tgt Ion: 65 Resp: 79258
Ion Ratio Lower Upper
65 100
67 53.1 0.0 99.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.691 min Scan# 1127
Delta R.T. 0.006 min
Lab File: VY008929.D
Acq: 21 May 2022 02:02




Instrument :
MSVOA_Y
ClientSampleId :
PP-1


Tgt Ion:114 Resp: 230210
Ion Ratio Lower Upper
114 100
63 19.1 0.0 43.2
88 13.4 0.0 31.8



#35
Dibromofluoromethane
Concen: 50.115 ug/l
RT: 7.716 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY008929.D
Acq: 21 May 2022 02:02

Tgt Ion:113 Resp: 71741
Ion Ratio Lower Upper
113 100
111 104.0 82.6 124.0
192 19.1 17.6 26.4









#50
Toluene-d8
Concen: 54.332 ug/l
RT: 10.179 min Scan# 1
Delta R.T. -0.000 min
Lab File: VY008929.D
Acq: 21 May 2022 02:02




Instrument :
MSVOA_Y
ClientSampleId :
PP-1

Tgt Ion: 98 Resp: 286938
Ion Ratio Lower Upper
98 100
100 64.9 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 42.602 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008929.D
Acq: 21 May 2022 02:02

Tgt Ion: 95 Resp: 90758
Ion Ratio Lower Upper
95 100
174 90.4 0.0 177.0
176 86.9 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 117
Delta R.T. -0.001 min
Lab File: VY008929.D
Acq: 21 May 2022 02:02

Instrument :
MSVOA_Y
ClientSampleId :
PP-1

Tgt Ion:117 Resp: 215060
Ion Ratio Lower Upper
117 100
82 55.3 43.0 64.4
119 31.2 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008929.D
Acq: 21 May 2022 02:02

Tgt Ion:152 Resp: 93383
Ion Ratio Lower Upper
152 100
115 54.7 28.2 84.7
150 153.8 0.0 347.6

