

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052323\
 Data File : VY013834.D
 Acq On : 23 May 2023 13:38
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: May 24 05:29:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.789	168	242602	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	385212	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	338181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	156205	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	155629	55.450	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.900%	
35) Dibromofluoromethane	7.716	113	132463	53.247	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.500%	
50) Toluene-d8	10.179	98	462620	48.928	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 97.860%	
62) 4-Bromofluorobenzene	12.477	95	148452	47.235	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 94.460%	
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	1950	2.341	ug/l	99
20) Methylene Chloride	4.722	84	18506	3.925	ug/l	96
89) n-Butylbenzene	13.696	91	10410	1.058	ug/l	99
93) 1,2,4-Trichlorobenzene	15.007	180	6282	2.069	ug/l	99
94) Hexachlorobutadiene	15.111	225	3502	1.830	ug/l	97
95) Naphthalene	15.227	128	11346	6.096	ug/l	99
96) 1,2,3-Trichlorobenzene	15.422	180	6582	2.438	ug/l	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052323\
Data File : VY013834.D
Acq On : 23 May 2023 13:38
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: May 24 05:29:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY013834.D
Acq: 23 May 2023 13:38




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion:168 Resp: 242602
Ion Ratio Lower Upper
168 100
99 53.5 43.4 65.2



#16
Acetone
Concen: 2.341 ug/l
RT: 3.960 min Scan# 351
Delta R.T. -0.000 min
Lab File: VY013834.D
Acq: 23 May 2023 13:38

Tgt Ion: 43 Resp: 1950
Ion Ratio Lower Upper
43 100
58 29.6 23.3 34.9









#20
Methylene Chloride
Concen: 3.925 ug/l
RT: 4.722 min Scan# 41
Delta R.T. -0.006 min
Lab File: VY013834.D
Acq: 23 May 2023 13:38




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion:	84	Resp:	18506
Ion	Ratio	Lower	Upper
84	100		
49	128.1	106.2	159.2
51	43.9	31.4	47.0
86	64.0	50.6	75.8



#33
1,2-Dichloroethane-d4
Concen: 55.450 ug/l
RT: 8.142 min Scan# 1037
Delta R.T. -0.006 min
Lab File: VY013834.D
Acq: 23 May 2023 13:38

Tgt Ion:	65	Resp:	155629
Ion	Ratio	Lower	Upper
65	100		
67	51.7	0.0	104.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1128

Delta R.T. -0.006 min

Lab File: VY013834.D

Acq: 23 May 2023 13:38

Instrument :

MSVOA_Y

ClientSampleId :

VIBLK





Tgt Ion:114 Resp: 385212

Ion Ratio Lower Upper

114 100

63 20.9 0.0 40.2

88 15.2 0.0 30.0



#35

Dibromofluoromethane

Concen: 53.247 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.006 min

Lab File: VY013834.D

Acq: 23 May 2023 13:38




Tgt Ion:113 Resp: 132463


Ion Ratio Lower Upper

113 100

111 102.0 80.4 120.6

192 21.0 16.6 25.0









#50
Toluene-d8
Concen: 48.928 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY013834.D
Acq: 23 May 2023 13:38




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion: 98 Resp: 462620
Ion Ratio Lower Upper
98 100
100 62.2 50.8 76.2



#62
4-Bromofluorobenzene
Concen: 47.235 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY013834.D
Acq: 23 May 2023 13:38

Tgt Ion: 95 Resp: 148452
Ion Ratio Lower Upper
95 100
174 88.0 0.0 176.0
176 85.2 0.0 170.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY013834.D

Acq: 23 May 2023 13:38

Instrument :

MSVOA_Y

ClientSampleId :

VIBLK





Tgt Ion:117 Resp: 338181

Ion Ratio Lower Upper

117 100

82 54.4 44.2 66.4

119 31.3 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. -0.006 min

Lab File: VY013834.D

Acq: 23 May 2023 13:38




Tgt Ion:152 Resp: 156205


Ion Ratio Lower Upper

152 100

115 57.5 28.8 86.5

150 156.1 0.0 353.4





#89

n-Butylbenzene

Concen: 1.058 ug/l

RT: 13.696 min Scan# 1948

Delta R.T. -0.000 min

Lab File: VY013834.D

Acq: 23 May 2023 13:38

Instrument :

MSVOA_Y

ClientSampleId :

VIBLK





Tgt Ion: 91 Resp: 10410

Ion Ratio Lower Upper

91 100

92 50.4 25.6 76.6

134 26.5 13.0 38.9



#93

1,2,4-Trichlorobenzene

Concen: 2.069 ug/l

RT: 15.007 min Scan# 2163

Delta R.T. -0.000 min

Lab File: VY013834.D

Acq: 23 May 2023 13:38




Tgt Ion: 180 Resp: 6282


Ion Ratio Lower Upper

180 100

182 96.1 48.1 144.3

145 28.4 14.8 44.4









#94
Hexachlorobutadiene
Concen: 1.830 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. -0.000 min
Lab File: VY013834.D
Acq: 23 May 2023 13:38




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion:225 Resp: 3502
Ion Ratio Lower Upper
225 100
223 64.9 31.3 93.9
227 63.6 32.9 98.6



#95
Naphthalene
Concen: 6.096 ug/l
RT: 15.227 min Scan# 2199
Delta R.T. -0.012 min
Lab File: VY013834.D
Acq: 23 May 2023 13:38


Tgt Ion:128 Resp: 11346
Ion Ratio Lower Upper
128 100
127 14.3 10.7 16.1
129 11.1 9.0 13.4





#96
 1,2,3-Trichlorobenzene
 Concen: 2.438 ug/l
 RT: 15.422 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VY013834.D
 Acq: 23 May 2023 13:38

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK



Tgt Ion:	180	Resp:	6582
Ion	Ratio	Lower	Upper
180	100		
182	101.6	47.4	142.2
145	31.8	15.5	46.5

