

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
 Data File : VY013844.D
 Acq On : 24 May 2023 14:25
 Operator : KP/MD
 Sample : 02899-04
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB01

Quant Time: May 25 07:08:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	181386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	304437	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.490	117	291603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	113816	54.238	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.480%	
35) Dibromofluoromethane	7.716	113	101756	51.756	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.520%	
50) Toluene-d8	10.179	98	373355	49.964	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.920%	
62) 4-Bromofluorobenzene	12.477	95	113595	45.734	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 91.460%	
Target Compounds						Qvalue
20) Methylene Chloride	4.722	84	20986	7.336	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
Data File : VY013844.D
Acq On : 24 May 2023 14:25
Operator : KP/MD
Sample : 02899-04
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB01

Quant Time: May 25 07:08:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

