

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
 Data File : VY013847.D
 Acq On : 24 May 2023 15:35
 Operator : KP/MD
 Sample : 02899-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB06A

Quant Time: May 25 07:08:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	43033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	66631	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.489	117	59715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	22334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	25722	51.666	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.340%	
35) Dibromofluoromethane	7.716	113	23161	53.824	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.640%	
50) Toluene-d8	10.179	98	78353	47.909	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 95.820%	
62) 4-Bromofluorobenzene	12.477	95	20760	38.188	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 76.380%	
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	2439	16.506	ug/l #	86
20) Methylene Chloride	4.704	84	5010	7.399	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
Data File : VY013847.D
Acq On : 24 May 2023 15:35
Operator : KP/MD
Sample : 02899-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB06A

Quant Time: May 25 07:08:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

