

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
 Data File : VY013852.D
 Acq On : 24 May 2023 17:32
 Operator : KP/MD
 Sample : 02848-20
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP6

Quant Time: May 25 07:09:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	166367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	287650	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	285087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	132347	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	110755	57.544	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.080%
35) Dibromofluoromethane	7.716	113	97176	52.311	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.620%
50) Toluene-d8	10.179	98	354076	50.150	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.300%
62) 4-Bromofluorobenzene	12.477	95	115926	49.396	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
Data File : VY013852.D
Acq On : 24 May 2023 17:32
Operator : KP/MD
Sample : 02848-20
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP6

Quant Time: May 25 07:09:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

