

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
 Data File : VY013853.D
 Acq On : 24 May 2023 17:55
 Operator : KP/MD
 Sample : 02886-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: May 25 07:09:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	4590	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	7969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	8530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	4532	85.346	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 170.700%#	
35) Dibromofluoromethane	7.709	113	3513	68.261	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 136.520%	
50) Toluene-d8	10.178	98	9859	50.404	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.800%	
62) 4-Bromofluorobenzene	12.477	95	3072	47.249	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 94.500%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	1568	26.885	ug/l	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
Data File : VY013853.D
Acq On : 24 May 2023 17:55
Operator : KP/MD
Sample : 02886-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: May 25 07:09:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

