

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
 Data File : VY013858.D
 Acq On : 24 May 2023 19:52
 Operator : KP/MD
 Sample : 02908-07
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB13

Quant Time: May 25 07:10:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	59828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	88330	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	78048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	30560	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	30544	44.129	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.260%
35) Dibromofluoromethane	7.722	113	29748	52.149	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.300%
50) Toluene-d8	10.179	98	105047	48.452	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.900%
62) 4-Bromofluorobenzene	12.483	95	27352	37.954	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.900%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.722	84	7243	7.800	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052423\
Data File : VY013858.D
Acq On : 24 May 2023 19:52
Operator : KP/MD
Sample : 02908-07
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB13

Quant Time: May 25 07:10:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

