

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052623\
 Data File : VY013899.D
 Acq On : 26 May 2023 17:33
 Operator : KP/MD
 Sample : 02945-04
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB17B

Quant Time: May 29 04:35:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	32793	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	48755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	45155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	16848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	21034	55.443	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.880%	
35) Dibromofluoromethane	7.716	113	17188	54.589	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.180%	
50) Toluene-d8	10.179	98	56725	47.402	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 94.800%	
62) 4-Bromofluorobenzene	12.477	95	15423	38.773	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 77.540%	
Target Compounds						
					Qvalue	
16) Acetone	3.942	43	5191	46.101	ug/l	# 82
20) Methylene Chloride	4.722	84	7106	16.075	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052623\
Data File : VY013899.D
Acq On : 26 May 2023 17:33
Operator : KP/MD
Sample : 02945-04
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB17B

Quant Time: May 29 04:35:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

