

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052623\
 Data File : VY013900.D
 Acq On : 26 May 2023 17:57
 Operator : KP/MD
 Sample : 02945-06
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB23

Quant Time: May 29 04:35:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	55921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	90083	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	89435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	37236	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	40457	62.535	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	125.060%	
35) Dibromofluoromethane	7.716	113	32891	56.537	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	113.080%	
50) Toluene-d8	10.179	98	110685	50.059	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	100.120%	
62) 4-Bromofluorobenzene	12.477	95	33704	45.858	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	91.720%	
Target Compounds						
20) Methylene Chloride	4.710	84	10202	13.111	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052623\
Data File : VY013900.D
Acq On : 26 May 2023 17:57
Operator : KP/MD
Sample : 02945-06
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB23

Quant Time: May 29 04:35:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

