

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052623\
 Data File : VY013904.D
 Acq On : 26 May 2023 19:30
 Operator : KP/MD
 Sample : 02955-06
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB26

Quant Time: May 29 04:36:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	21293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	32390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	29505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	11350	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	15266	61.971	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.940%
35) Dibromofluoromethane	7.716	113	11864	56.718	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.440%
50) Toluene-d8	10.179	98	39073	49.148	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.300%
62) 4-Bromofluorobenzene	12.483	95	11303	42.772	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.540%
Target Compounds						
						Qvalue
16) Acetone	3.948	43	815	11.147	ug/l	# 78
20) Methylene Chloride	4.722	84	2896	9.093	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052623\
Data File : VY013904.D
Acq On : 26 May 2023 19:30
Operator : KP/MD
Sample : 02955-06
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB26

Quant Time: May 29 04:36:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

