

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052623\
 Data File : VY013905.D
 Acq On : 26 May 2023 19:54
 Operator : KP/MD
 Sample : 02955-08
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB22

Quant Time: May 29 04:37:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	30832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	42958	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	33884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	11979	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	14844	41.615	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	83.240%
35) Dibromofluoromethane	7.722	113	13708	49.412	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.820%
50) Toluene-d8	10.179	98	49695	47.131	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.260%
62) 4-Bromofluorobenzene	12.483	95	11098	31.665	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	63.320%
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	842	7.953	ug/l #	88
20) Methylene Chloride	4.716	84	2025	3.008	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052623\
Data File : VY013905.D
Acq On : 26 May 2023 19:54
Operator : KP/MD
Sample : 02955-08
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB22

Quant Time: May 29 04:37:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

