



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060622\
 Data File : VY009135.D
 Acq On : 06 Jun 2022 08:56
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Title : SW846 8260
 Last Update : Fri May 20 23:50:52 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.4	9092	PASS
75	95	30	60	58.1	21611	PASS
95	95	100	100	100.0	37187	PASS
96	95	5	9	6.6	2450	PASS
173	174	0.00	2	1.7	571	PASS
174	95	50	100	90.8	33784	PASS
175	174	5	9	7.7	2595	PASS
176	174	95	101	94.9	32072	PASS
177	176	5	9	6.3	2029	PASS