

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060822\
 Data File : VY009176.D
 Acq On : 08 Jun 2022 18:02
 Operator : KP/MD
 Sample : VSTDIC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 09 01:37:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060822S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 01:36:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	134187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	218651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	143827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	100899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	15390	11.898	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	23.800%#
35) Dibromofluoromethane	7.716	113	13383	10.815	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	21.640%#
50) Toluene-d8	10.173	98	51483	11.457	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	22.920%#
62) 4-Bromofluorobenzene	12.477	95	16707	8.738	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	17.480%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.900	85	13352	12.296	ug/l	99
3) Chloromethane	2.107	50	27011	12.400	ug/l	95
4) Vinyl Chloride	2.241	62	38754	11.798	ug/l	94
5) Bromomethane	2.638	94	35527	11.277	ug/l	95
6) Chloroethane	2.778	64	26711	11.261	ug/l	95
7) Trichlorofluoromethane	3.113	101	25373	10.901	ug/l	89
8) Diethyl Ether	3.521	74	7694	10.817	ug/l	72
9) 1,1,2-Trichlorotrifluo...	3.881	101	13650	10.782	ug/l	97
10) Methyl Iodide	4.088	142	11159	8.253	ug/l	96
11) Tert butyl alcohol	4.948	59	7109	48.584	ug/l #	92
12) 1,1-Dichloroethene	3.863	96	12495	10.327	ug/l	94
13) Acrolein	3.717	56	3976	54.269	ug/l	98
14) Allyl chloride	4.466	41	16947	9.630	ug/l #	79
15) Acrylonitrile	5.149	53	19593	49.366	ug/l	96
16) Acetone	3.942	43	15025	48.184	ug/l	99
17) Carbon Disulfide	4.180	76	40556	10.692	ug/l	97
18) Methyl Acetate	4.466	43	8851	8.302	ug/l #	87
19) Methyl tert-butyl Ether	5.204	73	36586	9.780	ug/l	99
20) Methylene Chloride	4.698	84	25389	13.492	ug/l #	79
21) trans-1,2-Dichloroethene	5.204	96	15288	10.243	ug/l #	80
22) Diisopropyl ether	6.112	45	42137	9.100	ug/l #	96
23) Vinyl Acetate	6.052	43	140080	46.902	ug/l #	95
24) 1,1-Dichloroethane	6.015	63	26627	10.099	ug/l	98
25) 2-Butanone	6.978	43	24457	46.256	ug/l #	89
26) 2,2-Dichloropropane	6.972	77	24267	9.590	ug/l	95
27) cis-1,2-Dichloroethene	6.978	96	17507	9.963	ug/l	84
28) Bromochloromethane	7.326	49	12350	10.146	ug/l #	96
29) Tetrahydrofuran	7.344	42	15852	44.638	ug/l #	83
30) Chloroform	7.502	83	29196	10.231	ug/l	98
31) Cyclohexane	7.783	56	25613	10.709	ug/l #	85
32) 1,1,1-Trichloroethane	7.698	97	24961	9.960	ug/l	99
36) 1,1-Dichloropropene	7.911	75	21563	10.124	ug/l	97
37) Ethyl Acetate	7.064	43	11844	9.708	ug/l	96
38) Carbon Tetrachloride	7.893	117	22714	10.220	ug/l	95
39) Methylcyclohexane	9.179	83	25116	9.854	ug/l	92
40) Benzene	8.155	78	61235	9.777	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060822\
 Data File : VY009176.D
 Acq On : 08 Jun 2022 18:02
 Operator : KP/MD
 Sample : VSTDIC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 09 01:37:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060822S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 01:36:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.307	41	6759m	9.996	ug/l	
42) 1,2-Dichloroethane	8.234	62	19740	10.281	ug/l	92
43) Isopropyl Acetate	8.271	43	21441	9.367	ug/l #	90
44) Trichloroethene	8.935	130	16643	10.367	ug/l	97
45) 1,2-Dichloropropane	9.209	63	15204	9.762	ug/l	95
46) Dibromomethane	9.295	93	10150	10.017	ug/l	92
47) Bromodichloromethane	9.490	83	22724	10.136	ug/l	94
48) Methyl methacrylate	9.289	41	9611	9.607	ug/l	88
49) 1,4-Dioxane	9.295	88	2687	200.243	ug/l #	82
51) 4-Methyl-2-Pentanone	10.069	43	58126	45.525	ug/l	96
52) Toluene	10.240	92	39993	9.525	ug/l	96
53) t-1,3-Dichloropropene	10.459	75	22331	9.313	ug/l	95
54) cis-1,3-Dichloropropene	9.923	75	25201	9.687	ug/l	90
55) 1,1,2-Trichloroethane	10.642	97	12538	9.114	ug/l	95
56) Ethyl methacrylate	10.508	69	14799	8.336	ug/l	91
57) 1,3-Dichloropropane	10.782	76	21127	9.304	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.783	63	41891	45.951	ug/l	96
59) 2-Hexanone	10.831	43	36454	41.539	ug/l	91
60) Dibromochloromethane	10.977	129	15494	9.205	ug/l	98
61) 1,2-Dibromoethane	11.087	107	12777	9.486	ug/l	99
64) Tetrachloroethene	10.715	164	14175	14.643	ug/l	95
65) Chlorobenzene	11.514	112	30086	9.757	ug/l	94
66) 1,1,1,2-Tetrachloroethane	11.581	131	10529	9.228	ug/l	96
67) Ethyl Benzene	11.587	91	48709	8.860	ug/l	98
68) m/p-Xylenes	11.697	106	38680	17.210	ug/l	87
69) o-Xylene	12.026	106	17888	8.398	ug/l	95
70) Styrene	12.044	104	32827	8.685	ug/l	95
71) Bromoform	12.203	173	9445	12.026	ug/l #	97
73) Isopropylbenzene	12.325	105	68547	8.267	ug/l	99
74) N-amyl acetate	12.142	43	18239	7.981	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.581	83	16152	8.957	ug/l	96
76) 1,2,3-Trichloropropane	12.630	75	11357m	8.803	ug/l	
77) Bromobenzene	12.605	156	16497	8.524	ug/l	97
78) n-propylbenzene	12.666	91	85439	8.373	ug/l	99
79) 2-Chlorotoluene	12.758	91	46310	8.345	ug/l	94
80) 1,3,5-Trimethylbenzene	12.813	105	56189	7.922	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.374	75	5236	7.844	ug/l	91
82) 4-Chlorotoluene	12.855	91	48964	8.057	ug/l	96
83) tert-Butylbenzene	13.075	119	50125	8.126	ug/l	92
84) 1,2,4-Trimethylbenzene	13.117	105	56942	7.973	ug/l	97
85) sec-Butylbenzene	13.251	105	78709	9.583	ug/l	98
86) p-Isopropyltoluene	13.367	119	61944	9.285	ug/l	98
87) 1,3-Dichlorobenzene	13.361	146	34421	9.621	ug/l	98
88) 1,4-Dichlorobenzene	13.447	146	35255	10.076	ug/l	99
89) n-Butylbenzene	13.696	91	59238	9.619	ug/l	99
90) Hexachloroethane	13.959	117	13552	11.357	ug/l	80
91) 1,2-Dichlorobenzene	13.739	146	30549	10.875	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.349	75	2404	11.089	ug/l	84
93) 1,2,4-Trichlorobenzene	15.007	180	17437	11.225	ug/l	97
94) Hexachlorobutadiene	15.111	225	10510	12.319	ug/l	98
95) Naphthalene	15.239	128	33883	10.745	ug/l	99
96) 1,2,3-Trichlorobenzene	15.422	180	15418	11.332	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060822\
Data File : VY009176.D
Acq On : 08 Jun 2022 18:02
Operator : KP/MD
Sample : VSTDIC010
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1


Quant Time: Jun 09 01:37:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060822S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 01:36:09 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060822\
Data File : VY009176.D
Acq On : 08 Jun 2022 18:02
Operator : KP/MD
Sample : VSTDIC010
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1





Quant Time: Jun 09 01:37:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060822S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 01:36:09 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 978
Delta R.T. -0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:168 Resp: 134187
Ion Ratio Lower Upper
168 100
99 51.9 46.9 70.3



#2
Dichlorodifluoromethane
Concen: 12.296 ug/l
RT: 1.900 min Scan# 13
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 85 Resp: 13352
Ion Ratio Lower Upper
85 100
87 31.5 16.1 48.2








#3
Chloromethane
Concen: 12.400 ug/l
RT: 2.107 min Scan# 47
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 50 Resp: 27011
Ion Ratio Lower Upper
50 100
52 33.6 24.6 37.0



#4
Vinyl Chloride
Concen: 11.798 ug/l
RT: 2.241 min Scan# 69
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 62 Resp: 38754
Ion Ratio Lower Upper
62 100
64 34.6 25.0 37.6








#5
Bromomethane
Concen: 11.277 ug/l
RT: 2.638 min Scan# 134
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 94 Resp: 35527
Ion Ratio Lower Upper
94 100
96 87.5 73.5 110.3



#6
Chloroethane
Concen: 11.261 ug/l
RT: 2.778 min Scan# 157
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 64 Resp: 26711
Ion Ratio Lower Upper
64 100
66 34.7 25.4 38.0








#7
 Trichlorofluoromethane
 Concen: 10.901 ug/l
 RT: 3.113 min Scan# 212
 Delta R.T. 0.000 min
 Lab File: VY009176.D
 Acq: 08 Jun 2022 18:02


Tgt Ion:101 Resp: 25373
 Ion Ratio Lower Upper
 101 100
 103 68.2 48.0 72.0



#8
 Diethyl Ether
 Concen: 10.817 ug/l
 RT: 3.521 min Scan# 279
 Delta R.T. 0.000 min
 Lab File: VY009176.D
 Acq: 08 Jun 2022 18:02





Tgt Ion: 74 Resp: 7694
 Ion Ratio Lower Upper
 74 100
 45 81.0 55.1 165.5








#9
1,1,2-Trichlorotrifluoroethane
Concen: 10.782 ug/l
RT: 3.881 min Scan# 338
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:101 Resp: 13650
Ion Ratio Lower Upper
101 100
85 48.2 40.6 60.8
151 78.5 64.9 97.3



#10
Methyl Iodide
Concen: 8.253 ug/l
RT: 4.088 min Scan# 372
Delta R.T. 0.012 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion:142 Resp: 11159
Ion Ratio Lower Upper
142 100
127 48.8 41.5 62.3
141 15.0 11.9 17.9









#11

Tert butyl alcohol
Concen: 48.584 ug/l
RT: 4.948 min Scan# 513
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02




Tgt Ion: 59 Resp: 7109
Ion Ratio Lower Upper
59 100
57 5.0 6.4 9.6#




#12

1,1-Dichloroethene
Concen: 10.327 ug/l
RT: 3.863 min Scan# 335
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 96 Resp: 12495
Ion Ratio Lower Upper
96 100
61 176.9 145.8 218.6
98 75.6 52.5 78.7








#13
Acrolein
Concen: 54.269 ug/l
RT: 3.717 min Scan# 311
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 56 Resp: 3976
Ion Ratio Lower Upper
56 100
55 70.3 57.6 86.4



#14
Allyl chloride
Concen: 9.630 ug/l
RT: 4.466 min Scan# 434
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 41 Resp: 16947
Ion Ratio Lower Upper
41 100
39 90.0 58.8 88.2#
76 45.6 25.1 37.7#








#15
Acrylonitrile
Concen: 49.366 ug/l
RT: 5.149 min Scan# 546
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:	53	Resp:	19593
Ion	Ratio	Lower	Upper
53	100		
52	79.1	67.0	100.4
51	37.8	29.8	44.6



#16
Acetone
Concen: 48.184 ug/l
RT: 3.942 min Scan# 348
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:	43	Resp:	15025
Ion	Ratio	Lower	Upper
43	100		
58	27.0	22.0	33.0








#17
Carbon Disulfide
Concen: 10.692 ug/l
RT: 4.180 min Scan# 387
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 76 Resp: 40556
Ion Ratio Lower Upper
76 100
78 10.5 7.4 11.0



#18
Methyl Acetate
Concen: 8.302 ug/l
RT: 4.466 min Scan# 434
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 43 Resp: 8851
Ion Ratio Lower Upper
43 100
74 26.8 16.6 24.8#








#19
Methyl tert-butyl Ether
Concen: 9.780 ug/l
RT: 5.204 min Scan# 555
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 73 Resp: 36586
Ion Ratio Lower Upper
73 100
57 23.5 18.6 27.8



#20
Methylene Chloride
Concen: 13.492 ug/l
RT: 4.698 min Scan# 472
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion: 84 Resp: 25389
Ion Ratio Lower Upper
84 100
49 107.2 114.4 171.6#
51 33.0 35.3 52.9#
86 61.2 53.7 80.5









#21
trans-1,2-Dichloroethene
Concen: 10.243 ug/l
RT: 5.204 min Scan# 555
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion: 96 Resp: 15288




Ion	Ratio	Lower	Upper
96	100		
61	115.8	120.5	180.7#
98	64.1	49.8	74.8




#22
Diisopropyl ether
Concen: 9.100 ug/l
RT: 6.112 min Scan# 704
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion: 45 Resp: 42137





Ion	Ratio	Lower	Upper
45	100		
43	62.7	49.3	73.9
87	27.6	19.2	28.8
59	13.5	8.4	12.6#








#23
 Vinyl Acetate
 Concen: 46.902 ug/l
 RT: 6.052 min Scan# 694
 Delta R.T. 0.000 min
 Lab File: VY009176.D
 Acq: 08 Jun 2022 18:02


Tgt Ion: 43 Resp: 140080
 Ion Ratio Lower Upper
 43 100
 86 10.8 7.1 10.7#



#24
 1,1-Dichloroethane
 Concen: 10.099 ug/l
 RT: 6.015 min Scan# 688
 Delta R.T. 0.006 min
 Lab File: VY009176.D
 Acq: 08 Jun 2022 18:02

Tgt Ion: 63 Resp: 26627
 Ion Ratio Lower Upper
 63 100
 98 8.3 3.6 10.8
 100 3.8 2.0 5.8





#25

2-Butanone

Concen: 46.256 ug/l

RT: 6.978 min Scan# 846

Delta R.T. 0.000 min

Lab File: VY009176.D





Acq: 08 Jun 2022 18:02

Tgt Ion: 43 Resp: 24457

Ion Ratio Lower Upper

43 100

72 27.5 17.7 26.5#



#26

2,2-Dichloropropane

Concen: 9.590 ug/l

RT: 6.972 min Scan# 845

Delta R.T. 0.000 min

Lab File: VY009176.D




Acq: 08 Jun 2022 18:02


Tgt Ion: 77 Resp: 24267

Ion Ratio Lower Upper

77 100

97 23.3 10.4 31.4





#27

cis-1,2-Dichloroethene

Concen: 9.963 ug/l

RT: 6.978 min Scan# 846





Delta R.T. 0.000 min

Lab File: VY009176.D

Acq: 08 Jun 2022 18:02

Tgt Ion: 96 Resp: 17507

Ion	Ratio	Lower	Upper
96	100		
61	131.9	0.0	319.8
98	60.8	0.0	127.0



#28

Bromochloromethane

Concen: 10.146 ug/l

RT: 7.326 min Scan# 903




Delta R.T. 0.000 min


Lab File: VY009176.D

Acq: 08 Jun 2022 18:02

Tgt Ion: 49 Resp: 12350

Ion	Ratio	Lower	Upper
49	100		
129	2.6	0.0	2.2#
130	72.3	55.4	83.0





#29

Tetrahydrofuran

Concen: 44.638 ug/l





RT: 7.344 min Scan# 906

Delta R.T. 0.006 min

Lab File: VY009176.D

Acq: 08 Jun 2022 18:02

Tgt Ion:	42	Resp:	15852
Ion	Ratio	Lower	Upper
42	100		
72	47.8	30.3	45.5#
71	47.1	29.0	43.4#



#30

Chloroform

Concen: 10.231 ug/l




RT: 7.502 min Scan# 932


Delta R.T. 0.006 min

Lab File: VY009176.D

Acq: 08 Jun 2022 18:02





Tgt Ion:	83	Resp:	29196
Ion	Ratio	Lower	Upper
83	100		
85	62.8	51.8	77.6








#31
Cyclohexane
Concen: 10.709 ug/l
RT: 7.783 min Scan# 978
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 56 Resp: 25613
Ion Ratio Lower Upper
56 100
69 40.8 26.7 40.1#
84 98.1 67.2 100.8



#32
1,1,1-Trichloroethane
Concen: 9.960 ug/l
RT: 7.698 min Scan# 964
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 97 Resp: 24961
Ion Ratio Lower Upper
97 100
99 62.8 50.7 76.1
61 45.5 37.6 56.4








#33
1,2-Dichloroethane-d4
Concen: 11.898 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 65 Resp: 15390
Ion Ratio Lower Upper
65 100
67 51.9 0.0 99.4



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.685 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 114 Resp: 218651
Ion Ratio Lower Upper
114 100
63 18.2 0.0 43.2
88 12.3 0.0 31.8








#35
Dibromofluoromethane
Concen: 10.815 ug/l
RT: 7.716 min Scan# 967
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:113 Resp: 13383
Ion Ratio Lower Upper
113 100
111 106.1 82.6 124.0
192 19.2 17.6 26.4



#36
1,1-Dichloropropene
Concen: 10.124 ug/l
RT: 7.911 min Scan# 999
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion: 75 Resp: 21563
Ion Ratio Lower Upper
75 100
110 36.4 17.1 51.2
77 31.9 24.4 36.6









#37
Ethyl Acetate
Concen: 9.708 ug/l
RT: 7.064 min Scan# 860
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion: 43 Resp: 11844




Ion	Ratio	Lower	Upper
43	100		
61	14.2	9.7	14.5
70	9.4	7.1	10.7




#38
Carbon Tetrachloride
Concen: 10.220 ug/l
RT: 7.893 min Scan# 996
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion: 117 Resp: 22714

Ion	Ratio	Lower	Upper
117	100		
119	103.2	77.9	116.9
121	32.8	24.8	37.2





#39

Methylcyclohexane

Concen: 9.854 ug/l





RT: 9.179 min Scan# 1207

Delta R.T. 0.000 min

Lab File: VY009176.D

Acq: 08 Jun 2022 18:02

Tgt Ion:	83	Resp:	25116
Ion	Ratio	Lower	Upper
83	100		
55	72.0	66.5	99.7
98	46.7	37.9	56.9



#40

Benzene

Concen: 9.777 ug/l




RT: 8.155 min Scan# 1039


Delta R.T. 0.000 min

Lab File: VY009176.D

Acq: 08 Jun 2022 18:02





Tgt Ion:	78	Resp:	61235
Ion	Ratio	Lower	Upper
78	100		
77	23.1	20.4	30.6








#41
Methacrylonitrile
Concen: 9.996 ug/l m
RT: 7.307 min Scan# 900
Delta R.T. -0.012 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 41 Resp: 6759
Ion Ratio Lower Upper
41 100
39 146.8 92.9 139.3#
67 0.0 0.0 0.0
52 0.0 0.0 0.0



#42
1,2-Dichloroethane
Concen: 10.281 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 62 Resp: 19740
Ion Ratio Lower Upper
62 100
98 10.1 0.0 14.8








#43
Isopropyl Acetate
Concen: 9.367 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 43 Resp: 21441
Ion Ratio Lower Upper
43 100
61 20.0 20.2 30.2#
87 14.9 8.9 13.3#



#44
Trichloroethene
Concen: 10.367 ug/l
RT: 8.935 min Scan# 1167
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion:130 Resp: 16643
Ion Ratio Lower Upper
130 100
95 91.2 0.0 188.2





#45

1,2-Dichloropropane

Concen: 9.762 ug/l

RT: 9.209 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VY009176.D





Acq: 08 Jun 2022 18:02

Tgt Ion: 63 Resp: 15204

Ion Ratio Lower Upper

63 100

65 31.2 23.0 34.6



#46

Dibromomethane

Concen: 10.017 ug/l

RT: 9.295 min Scan# 1226

Delta R.T. -0.006 min

Lab File: VY009176.D

Acq: 08 Jun 2022 18:02




Tgt Ion: 93 Resp: 10150


Ion Ratio Lower Upper

93 100

95 78.3 64.3 96.5





174 91.2 83.4 125.0








#47
Bromodichloromethane
Concen: 10.136 ug/l
RT: 9.490 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:	83	Resp:	22724
Ion	Ratio	Lower	Upper
83	100		
85	59.8	51.9	77.9
127	8.2	6.6	10.0



#48
Methyl methacrylate
Concen: 9.607 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:	41	Resp:	9611
Ion	Ratio	Lower	Upper
41	100		
69	83.7	58.2	87.2
39	66.4	47.4	71.0








#49
1,4-Dioxane
Concen: 200.243 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 88 Resp: 2687
Ion Ratio Lower Upper
88 100
43 31.5 32.8 49.2#
58 57.6 59.8 89.6#



#50
Toluene-d8
Concen: 11.457 ug/l
RT: 10.173 min Scan# 1370
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 98 Resp: 51483
Ion Ratio Lower Upper
98 100
100 62.8 50.2 75.2








#51
4-Methyl-2-Pentanone
Concen: 45.525 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 43 Resp: 58126
Ion Ratio Lower Upper
43 100
58 37.4 28.1 42.1



#52
Toluene
Concen: 9.525 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 92 Resp: 39993
Ion Ratio Lower Upper
92 100
91 172.0 142.6 213.8








#53
t-1,3-Dichloropropene
Concen: 9.313 ug/l
RT: 10.459 min Scan# 1417
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 75 Resp: 22331
Ion Ratio Lower Upper
75 100
77 28.9 25.4 38.2



#54
cis-1,3-Dichloropropene
Concen: 9.687 ug/l
RT: 9.923 min Scan# 1329
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 75 Resp: 25201
Ion Ratio Lower Upper
75 100
77 31.6 25.6 38.4
39 48.7 47.6 71.4








#55
1,1,2-Trichloroethane
Concen: 9.114 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:	97	Resp:	12538
Ion	Ratio	Lower	Upper
97	100		
83	83.4	69.8	104.8
85	61.4	44.3	66.5
99	64.1	49.8	74.6



#56
Ethyl methacrylate
Concen: 8.336 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:	69	Resp:	14799
Ion	Ratio	Lower	Upper
69	100		
41	66.6	63.4	95.0
39	44.9	35.8	53.8








#57
1,3-Dichloropropane
Concen: 9.304 ug/l
RT: 10.782 min Scan# 1470
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 76 Resp: 21127
Ion Ratio Lower Upper
76 100
78 32.9 25.0 37.6



#58
2-Chloroethyl Vinyl ether
Concen: 45.951 ug/l
RT: 9.783 min Scan# 1306
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 63 Resp: 41891
Ion Ratio Lower Upper
63 100
106 28.9 21.5 32.3








#59
2-Hexanone
Concen: 41.539 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 43 Resp: 36454
Ion Ratio Lower Upper
43 100
58 55.2 24.4 73.4



#60
Dibromochloromethane
Concen: 9.205 ug/l
RT: 10.977 min Scan# 1502
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 129 Resp: 15494
Ion Ratio Lower Upper
129 100
127 75.9 38.7 116.1








#61
1,2-Dibromoethane
Concen: 9.486 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 107 Resp: 12777
Ion Ratio Lower Upper
107 100
109 93.0 75.4 113.2



#62
4-Bromofluorobenzene
Concen: 8.738 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 95 Resp: 16707
Ion Ratio Lower Upper
95 100
174 85.9 0.0 177.0
176 83.2 0.0 175.8








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 1586
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:117 Resp: 143827
Ion Ratio Lower Upper
117 100
82 52.5 43.0 64.4
119 31.3 25.4 38.0



#64
Tetrachloroethene
Concen: 14.643 ug/l
RT: 10.715 min Scan# 1459
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:164 Resp: 14175
Ion Ratio Lower Upper
164 100
166 118.2 100.4 150.6
129 90.7 71.4 107.0
131 84.0 70.8 106.2








#65
Chlorobenzene
Concen: 9.757 ug/l
RT: 11.514 min Scan# 1590
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:112 Resp: 30086
Ion Ratio Lower Upper
112 100
114 28.9 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 9.228 ug/l
RT: 11.581 min Scan# 1601
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:131 Resp: 10529
Ion Ratio Lower Upper
131 100
133 94.5 47.9 143.6
119 68.0 30.8 92.3








#67
Ethyl Benzene
Concen: 8.860 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 91 Resp: 48709
Ion Ratio Lower Upper
91 100
106 31.7 24.5 36.7



#68
m/p-Xylenes
Concen: 17.210 ug/l
RT: 11.697 min Scan# 1620
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:106 Resp: 38680
Ion Ratio Lower Upper
106 100
91 188.5 166.6 249.8








#69
o-Xylene
Concen: 8.398 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:106 Resp: 17888
Ion Ratio Lower Upper
106 100
91 213.4 110.3 331.0



#70
Styrene
Concen: 8.685 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:104 Resp: 32827
Ion Ratio Lower Upper
104 100
78 47.6 42.4 63.6
103 53.2 44.4 66.6








#71
Bromoform
Concen: 12.026 ug/l
RT: 12.203 min Scan# 1703
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:	173	Resp:	9445
Ion Ratio	Lower	Upper	
173	100		
175	47.4	24.6	74.0
254	0.0	0.0	0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:	152	Resp:	100899
Ion Ratio	Lower	Upper	
152	100		
115	57.0	28.2	84.7
150	162.5	0.0	347.6








#73
Isopropylbenzene
Concen: 8.267 ug/l
RT: 12.325 min Scan# 1723
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:105 Resp: 68547
Ion Ratio Lower Upper
105 100
120 27.5 13.4 40.2



#74
N-ethyl acetate
Concen: 7.981 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 43 Resp: 18239
Ion Ratio Lower Upper
43 100
70 47.2 28.7 43.1#
55 25.4 17.3 25.9
61 23.3 17.0 25.6








#75
1,1,2,2-Tetrachloroethane
Concen: 8.957 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 83 Resp: 16152
Ion Ratio Lower Upper
83 100
131 9.4 5.6 16.8
85 67.3 32.3 96.8



#76
1,2,3-Trichloropropane
Concen: 8.803 ug/l m
RT: 12.630 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 75 Resp: 11357
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0








#77
Bromobenzene
Concen: 8.524 ug/l
RT: 12.605 min Scan# 1769
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:156 Resp: 16497
Ion Ratio Lower Upper
156 100
77 186.0 95.5 286.4
158 97.5 48.0 144.0



#78
n-propylbenzene
Concen: 8.373 ug/l
RT: 12.666 min Scan# 1779
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 91 Resp: 85439
Ion Ratio Lower Upper
91 100
120 23.1 11.3 33.9








#79
2-Chlorotoluene
Concen: 8.345 ug/l
RT: 12.758 min Scan# 1794
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 91 Resp: 46310
Ion Ratio Lower Upper
91 100
126 36.9 16.9 50.6



#80
1,3,5-Trimethylbenzene
Concen: 7.922 ug/l
RT: 12.813 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 105 Resp: 56189
Ion Ratio Lower Upper
105 100
120 50.7 24.4 73.4








#81
trans-1,4-Dichloro-2-butene
Concen: 7.844 ug/l
RT: 12.374 min Scan# 1731
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 75 Resp: 5236
Ion Ratio Lower Upper
75 100
53 93.2 85.3 127.9
89 46.8 37.0 55.4



#82
4-Chlorotoluene
Concen: 8.057 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 91 Resp: 48964
Ion Ratio Lower Upper
91 100
126 36.3 17.0 51.0








#83
tert-Butylbenzene
Concen: 8.126 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:119 Resp: 50125
Ion Ratio Lower Upper
119 100
91 61.2 34.8 104.4
134 26.7 13.3 39.9



#84
1,2,4-Trimethylbenzene
Concen: 7.973 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:105 Resp: 56942
Ion Ratio Lower Upper
105 100
120 46.3 22.3 66.9








#85
 sec-Butylbenzene
 Concen: 9.583 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VY009176.D
 Acq: 08 Jun 2022 18:02


Tgt Ion:105 Resp: 78709
 Ion Ratio Lower Upper
 105 100
 134 19.2 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 9.285 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY009176.D
 Acq: 08 Jun 2022 18:02





Tgt Ion:119 Resp: 61944
 Ion Ratio Lower Upper
 119 100
 134 27.7 13.2 39.5
 91 24.3 12.4 37.2








#87
1,3-Dichlorobenzene
Concen: 9.621 ug/l
RT: 13.361 min Scan# 1893
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:146 Resp: 34421
Ion Ratio Lower Upper
146 100
111 37.3 19.5 58.5
148 64.8 31.6 95.0



#88
1,4-Dichlorobenzene
Concen: 10.076 ug/l
RT: 13.447 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:146 Resp: 35255
Ion Ratio Lower Upper
146 100
111 40.3 19.3 57.8
148 65.4 32.6 97.8








#89
n-Butylbenzene
Concen: 9.619 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion: 91 Resp: 59238
Ion Ratio Lower Upper
91 100
92 52.1 26.3 78.9
134 24.4 13.0 38.9



#90
Hexachloroethane
Concen: 11.357 ug/l
RT: 13.959 min Scan# 1991
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion: 117 Resp: 13552
Ion Ratio Lower Upper
117 100
201 67.2 43.0 128.8








#91
1,2-Dichlorobenzene
Concen: 10.875 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion	Ratio	Lower	Upper
146	100		
111	40.2	20.3	60.9
148	63.4	32.3	96.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 11.089 ug/l
RT: 14.349 min Scan# 2055
Delta R.T. -0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion	Ratio	Lower	Upper
75	100		
155	91.7	38.0	113.9
157	108.4	47.8	143.3








#93
1,2,4-Trichlorobenzene
Concen: 11.225 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02


Tgt Ion:180 Resp: 17437
Ion Ratio Lower Upper
180 100
182 93.9 47.3 141.9
145 24.5 14.6 43.8



#94
Hexachlorobutadiene
Concen: 12.319 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02





Tgt Ion:225 Resp: 10510
Ion Ratio Lower Upper
225 100
223 64.8 30.8 92.3
227 62.8 31.6 94.7





#95
Naphthalene
Concen: 10.745 ug/l
RT: 15.239 min Scan# 2201
Delta R.T. 0.006 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion:128	Resp:	33883
Ion Ratio	Lower	Upper
128	100	
127	13.5	11.0 16.4
129	11.1	8.4 12.6



#96
1,2,3-Trichlorobenzene
Concen: 11.332 ug/l
RT: 15.422 min Scan# 2231
Delta R.T. 0.000 min
Lab File: VY009176.D
Acq: 08 Jun 2022 18:02

Tgt Ion:180	Resp:	15418
Ion Ratio	Lower	Upper
180	100	
182	91.9	47.3 141.9
145	31.0	15.3 45.9

