



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060823\
Data File : VY014072.D
Acq On : 08 Jun 2023 12:30
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Title : SW846 8260
Last Update : Fri Jun 02 07:00:14 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	18595	PASS
75	95	30	60	55.5	49405	PASS
95	95	100	100	100.0	88992	PASS
96	95	5	9	6.7	5924	PASS
173	174	0.00	2	1.8	1419	PASS
174	95	50	100	90.1	80149	PASS
175	174	5	9	7.8	6223	PASS
176	174	95	101	97.6	78216	PASS
177	176	5	9	6.5	5104	PASS