

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060923\
 Data File : VY014106.D
 Acq On : 09 Jun 2023 19:45
 Operator : KP/MD
 Sample : 03044-14RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-08-0-2.0RE

Quant Time: Jun 10 05:15:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	21036	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	32038	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	30336	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	12204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	13460	45.950	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.900%
35) Dibromofluoromethane	7.722	113	11548	50.826	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.660%
50) Toluene-d8	10.185	98	38034	44.776	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.560%
62) 4-Bromofluorobenzene	12.477	95	10877	36.047	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	72.100%
Target Compounds						
16) Acetone	3.954	43	2132	24.145	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060923\
Data File : VY014106.D
Acq On : 09 Jun 2023 19:45
Operator : KP/MD
Sample : 03044-14RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-08-0-2.0RE

Quant Time: Jun 10 05:15:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

